

LB129588


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

55225



A= 0.0000e+000

B= 1.8553e-004 slope

C= -5.3755e-002 y-intercept

Rho= 0.9993654

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.00	0.000	0.047	0.047	543	0.000	543					0/00
0.05	0.050					442	0				=
0.20	0.200	0.243	0.043	1602	0.0 %	1602					22
2.50	2.500	2.501	0.001	13771	0.0 %	13771					0
5.00	5.000	4.963	-0.037	27042	0.0 %	27042					-1
7.50	7.500	7.253	-0.247	39380	0.0 %	39380					-3
10.0	10.000	10.192	0.192	55225	0.0 %	55225					2

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	543	-		0 SFAM01.1	PPB	2/19/2024 15:56	Std
	0.2 S01	1602	-		0.2 SFAM01.1	PPB	2/19/2024 15:58	Std
	2.5 S02	13771	-		2.5 SFAM01.1	PPB	2/19/2024 16:01	Std
	5 S03	27042	-		5 SFAM01.1	PPB	2/19/2024 16:03	Std
	7.5 S04	39380	-		7.5 SFAM01.1	PPB	2/19/2024 16:16	Std
	10 S05	55225	-		10 SFAM01.1	PPB	2/19/2024 16:18	Std
ICV015	ICV015	21941	4.017 -		SFAM01.1	PPB	2/19/2024 16:21	SMPL
ICB015	ICB015	413	0.0229 -		SFAM01.1	PPB	2/19/2024 16:23	SMPL
CCV048	CCV048	26876	4.9326 -		SFAM01.1	PPB	2/19/2024 16:26	SMPL
CCB048	CCB048	391	0.0188 -		SFAM01.1	PPB	2/19/2024 16:28	SMPL
PB159121BL	PBS912	170	-0.0222 -		SFAM01.1	PPB	2/19/2024 16:33	SMPL
P1350-01	MBH6T7	4012	0.6906 -		SFAM01.1	PPB	2/19/2024 16:35	SMPL
P1350-02	MBH6W3	552	0.0487 -		SFAM01.1	PPB	2/19/2024 16:37	SMPL
P1350-03	MBH6W4	2132	0.3418 -		SFAM01.1	PPB	2/19/2024 16:40	SMPL
P1350-04	MBH6W5	2366	0.3852 -		SFAM01.1	PPB	2/19/2024 16:42	SMPL
P1350-05	MBH6X3	981	0.1283 -		SFAM01.1	PPB	2/19/2024 16:44	SMPL
P1350-06	MBH6X3D	966	0.1255 -		SFAM01.1	PPB	2/19/2024 16:46	SMPL
P1350-07	MBH6X3S	14314	2.602 -		SFAM01.1	PPB	2/19/2024 16:49	SMPL
P1351-01	MCOFN3	3844	0.6594 -		SFAM01.1	PPB	2/19/2024 16:51	SMPL
P1351-02	MCOFN3D	3739	0.64 -		SFAM01.1	PPB	2/19/2024 16:53	SMPL
P1351-03	MCOFN3S	15605	2.8415 -		SFAM01.1	PPB	2/19/2024 16:55	SMPL
P1351-04	MCOFN9	2807	0.467 -		SFAM01.1	PPB	2/19/2024 16:58	SMPL
P1351-05	MCOFPO	4633	0.8058 -		SFAM01.1	PPB	2/19/2024 17:00	SMPL
CCV049	CCV049	27132	4.9801 -		SFAM01.1	PPB	2/19/2024 17:02	SMPL
CCB049	CCB049	360	0.013 -		SFAM01.1	PPB	2/19/2024 17:05	SMPL