



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB129595

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 99.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/20/2024	10:25
2	CAL2	1	Water	NA	NA	20.3	7.01	02/20/2024	10:26
3	CAL3	1	Water	NA	NA	20.3	10.03	02/20/2024	10:29
4	ICV	1	Water	NA	NA	20.4	7.02	02/20/2024	10:33
5	CCV1	1	Water	NA	NA	20.2	2.01	02/20/2024	10:35
6	P1528-01	1	Water	NA	NA	22.2	7.33	02/20/2024	10:40
7	P1529-01	1	Water	NA	NA	22.1	7.36	02/20/2024	10:45
8	P1529-03	1	Water	NA	NA	22.1	7.77	02/20/2024	10:48
9	P1529-05	1	Water	NA	NA	22.1	8.04	02/20/2024	10:50
10	P1529-07	1	Water	NA	NA	21.9	7.91	02/20/2024	10:55
11	P1529-09	1	Water	NA	NA	22.1	8.10	02/20/2024	11:00
12	P1531-01	1	Water	NA	NA	20.7	8.19	02/20/2024	11:10
13	P1536-01	1	Water	NA	NA	20.1	5.36	02/20/2024	11:15
14	P1536-01DUP	1	Water	NA	NA	20.2	5.37	02/20/2024	11:16
15	CCV2	1	Water	NA	NA	20.3	12.02	02/20/2024	11:39