


LB129597

7471B

INSTRUMENT ID: CVI



Linear

A= 0.0000e+000

B= 1.7652e-004

C= -4.6775e-002

Rho= 0.9999246

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.031	0.031	440	0.000	440					
0.2	0.200	0.192	-0.008	1350	0.0 %	1350					-4
2.5	2.500	2.426	-0.074	14006	0.0 %	14006					-3
5.0	5.000	5.074	0.074	29011	0.0 %	29011					1
7.5	7.500	7.489	-0.011	42693	0.0 %	42693					0
10.0	10.000	9.988	-0.012	56850	0.0 %	56850					0

LB129597 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Feb 2024 12:44:02

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	440	-	0.0000	7471B	PPB	20 Feb 2024 13:15:48	S	Std
0.2 - 1	50.2	1350	-	0.2000	7471B	PPB	20 Feb 2024 13:18:06	S	Std
2.5 - 1	52.5	14006	-	2.5000	7471B	PPB	20 Feb 2024 13:20:23	S	Std
5.0 - 1	55	29011	-	5.0000	7471B	PPB	20 Feb 2024 13:22:41	S	Std
7.5 - 1	57.5	42693	-	7.5000	7471B	PPB	20 Feb 2024 13:25:03	S	Std
10.0 - 1	510	56850	-	10.0000	7471B	PPB	20 Feb 2024 13:30:16	S	Std
ICV17 - 1	ICV17	23233	4.0543	-	7471B	PPB	20 Feb 2024 13:33:07	U	SMPL
ICB17 - 1	ICB17	176	-0.0157	-	7471B	PPB	20 Feb 2024 13:35:25	U	SMPL
CCV55 - 1	CCV55	28970	5.0670	-	7471B	PPB	20 Feb 2024 13:37:50	U	SMPL
CCB55 - 1	CCB55	199	-0.0116	-	7471B	PPB	20 Feb 2024 13:40:05	U	SMPL
CRA - 1	CRA	1345	0.1906	-	7471B	PPB	20 Feb 2024 13:42:34	U	SMPL
HighStd - 1	HighStd	57355	10.0775	-	7471B	PPB	20 Feb 2024 13:44:50	U	SMPL
ChkStd - 1	ChkStd	40181	7.0460	-	7471B	PPB	20 Feb 2024 13:47:06	U	SMPL
PB159145BL - 1	PBS	110	-0.0274	-	7471B	PPB	20 Feb 2024 13:49:38	U	SMPL
PB159145BS - 1	LCSS	20148	3.5097	-	7471B	PPB	20 Feb 2024 13:52:09	U	SMPL
P1505-01 - 1	WC-1S	11320	1.9514	-	7471B	PPB	20 Feb 2024 13:59:18	U	SMPL
P1505-01DUP - 1	WC-1SDUP	11124	1.9168	-	7471B	PPB	20 Feb 2024 14:01:37	U	SMPL
P1505-01MS - 1	WC-1SMS	33216	5.8165	-	7471B	PPB	20 Feb 2024 14:03:57	U	SMPL
P1505-01MSD - 1	WC-1MSD	37520	6.5763	-	7471B	PPB	20 Feb 2024 14:06:17	U	SMPL
P1505-05 - 1	WC-1D	15450	2.6805	-	7471B	PPB	20 Feb 2024 14:08:43	U	SMPL
CCV56 - 1	CCV56	28996	5.0716	-	7471B	PPB	20 Feb 2024 14:11:13	U	SMPL
CCB56 - 1	CCB56	273	0.0014	-	7471B	PPB	20 Feb 2024 14:13:35	U	SMPL
P1505-09 - 1	WC-2S	7276	1.2376	-	7471B	PPB	20 Feb 2024 14:16:02	U	SMPL
P1505-13 - 1	WC-2D	41809	7.3333	-	7471B	PPB	20 Feb 2024 14:18:18	U	SMPL
P1514-01 - 1	ETGI-343	3307	0.5370	-	7471B	PPB	20 Feb 2024 14:20:37	U	SMPL
P1505-01LX5 - 1		2552	0.4037	-	7471B	PPB	20 Feb 2024 14:33:50	U	SMPL
P1505-01A - 1		35750	6.2638	-	7471B	PPB	20 Feb 2024 14:36:05	U	SMPL
CCV57 - 1	CCV57	29020	5.0758	-	7471B	PPB	20 Feb 2024 14:38:23	U	SMPL
CCB57 - 1	CCB57	231	-0.0060	-	7471B	PPB	20 Feb 2024 14:40:47	U	SMPL