



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Nikita

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129599

pH Meter ID: WC pH Meter-1

Reagent/Standard	Lot/Log #
hexavalent chromium color reagent	WP106629
5N sulfuric acid	WP105081
HNO3 Hex-Chrome, 5M	WP105080
Hexchrome Cleaning Solution	WP106152

Intercept: 0.001

Slope: 0.7633

Regression: 0.999998

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml)	Final Vol (ml)	pH HNO3	pH H2SO4	Absorb.at 540nm		Absorbance Difference	Result (mg/L)	%D	Anal Date	Anal Time
								Backgrnd	Color					
1	CAL1	0	1	100	100	7.85	1.85	0.000	0.000	0.000	-0.00		02/20/2024	11:45
2	CAL2	0.01	1	100	100	7.82	1.76	0.000	0.009	0.009	0.010	0	02/20/2024	11:46
3	CAL3	0.025	1	100	100	7.54	1.54	0.000	0.020	0.020	0.024	-4	02/20/2024	11:47
4	CAL4	0.05	1	100	100	7.83	1.76	0.000	0.039	0.039	0.049	-2	02/20/2024	11:48
5	CAL5	0.1	1	100	100	7.44	1.91	0.000	0.078	0.078	0.100	0	02/20/2024	11:49
6	CAL6	0.5	1	100	100	7.75	1.93	0.000	0.383	0.383	0.500	0	02/20/2024	11:50
7	CAL7	1	1	100	100	7.56	1.82	0.000	0.764	0.764	0.999	-0.1	02/20/2024	11:51



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Nikita

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129599

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb.at540nm		Absorbance Difference	Intermediate Result (mg/L)	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100	7.65	1.85	0.000	0.381	0.381	0.498	02/20/2024	11:52
2	ICB		1	100	100	7.66	1.85	0.000	0.001	0.001	0.000	02/20/2024	11:53
3	CCV1	0.5	1	100	100	7.28	1.76	0.000	0.385	0.385	0.503	02/20/2024	11:54
4	CCB1		1	100	100	7.45	1.91	0.000	0.001	0.001	0.000	02/20/2024	11:55
5	RL Check	0.01	1	100	100	7.54	1.93	0.000	0.009	0.009	0.010	02/20/2024	11:56
6	PB159131BL		1	2.50	100	7.44	1.82	0.000	0.000	0.000	-0.001	02/20/2024	11:57
7	PB159131BS	20	1	2.50	100	7.75	1.54	0.000	0.386	0.386	0.504	02/20/2024	11:58
8	P1454-03		1	2.55	100	7.54	2.21	0.026	0.028	0.002	0.001	02/20/2024	11:59
9	P1472-01		1	2.53	100	7.58	2.30	0.038	0.040	0.002	0.001	02/20/2024	12:00
10	P1523-01		1	2.52	100	7.58	2.19	0.009	0.010	0.001	0.000	02/20/2024	12:01
11	P1523-02		1	2.54	100	7.43	2.10	0.009	0.011	0.002	0.001	02/20/2024	12:02
12	P1523-03		1	2.55	100	7.43	2.16	0.008	0.010	0.002	0.001	02/20/2024	12:03
13	P1523-03DU		1	2.52	100	7.52	1.89	0.009	0.010	0.001	0.000	02/20/2024	12:04
14	P1523-03MS	40	2	2.51	100	7.74	1.89	0.003	0.322	0.319	0.417	02/20/2024	12:05
15	P1523-03MS	1284	40	2.52	100	7.54	2.00	0.000	0.597	0.597	0.781	02/20/2024	12:06
16	CCV2	0.5	1	100	100	7.43	2.10	0.000	0.385	0.385	0.503	02/20/2024	12:07
17	CCB2		1	100	100	7.33	2.09	0.000	0.001	0.001	0.000	02/20/2024	12:08
18	P1523-03MS	40	2	2.53	100	7.38	1.78	0.003	0.369	0.366	0.478	02/20/2024	12:09
19	P1523-06		1	2.51	100	7.38	2.10	0.006	0.008	0.002	0.001	02/20/2024	12:10
20	P1523-07		1	2.54	100	7.41	2.08	0.008	0.009	0.001	0.000	02/20/2024	12:11
21	P1523-08		1	2.53	100	7.75	2.30	0.009	0.010	0.001	0.000	02/20/2024	12:12
22	P1523-08DU		1	2.53	100	7.52	1.89	0.010	0.011	0.001	0.000	02/20/2024	12:13
23	P1523-08MS	40	2	2.54	100	7.82	1.89	0.005	0.317	0.312	0.407	02/20/2024	12:14
24	P1523-08MS	1284	40	2.54	100	7.72	2.00	0.000	0.592	0.592	0.774	02/20/2024	12:15
25	P1523-08MS	40	2	2.52	100	7.45	1.78	0.004	0.379	0.375	0.490	02/20/2024	12:16
26	CCV3	0.5	1	100	100	7.72	1.89	0.000	0.385	0.385	0.503	02/20/2024	12:17
27	CCB3		1	100	100	7.38	1.78	0.000	0.001	0.001	0.000	02/20/2024	12:18