



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129606

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	02/21/2024	08:15
2	P1516-02		1	14	100.00	>212.0	>212.0	02/21/2024	08:45
3	P1516-02DUP		1	12	100.00	>212.0	>212.0	02/21/2024	09:15
4	P1528-01		1	13	100.00	>212.0	>212.0	02/21/2024	09:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129606

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 02/21/24 flashpoint

WorkList ID : 178033

Department : Wet-Chemistry

Date : 02-21-2024 07:49:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1516-02	00042	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/16/2024	1010B
P1528-01	IDW-20240215	Water	Flash Point	Cool 4 deg C	ECOL03	E21	02/15/2024	1010B

Date/Time02/21/202408:05

Raw Sample Received by: N.P. W.C.

Raw Sample Relinquished by: [Signature]

Date/Time02/21/202410:15

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: N.P. W.C.