



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/22/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 02/21/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 02/22/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB129607

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1537-01	1	1	1.14	8.47	9.61	8.62	88.3	
P1537-02	2	2	1.18	8.63	9.81	8.77	87.9	
P1537-03	3	3	1.18	8.50	9.68	8.52	86.4	
P1537-04	4	4	1.19	8.52	9.71	8.61	87.1	
P1537-05	5	5	1.15	8.83	9.98	8.77	86.3	
P1537-06	6	6	1.16	8.80	9.96	8.86	87.5	
P1537-07	7	7	1.19	8.58	9.77	8.87	89.5	
P1537-08	8	8	1.18	8.52	9.7	8.44	85.2	
P1537-09	9	9	1.18	8.48	9.66	8.5	86.3	
P1537-10	10	10	1.15	8.40	9.55	8.34	85.6	
P1537-11	11	11	1.18	8.79	9.97	9.05	89.5	
P1537-12	12	12	1.18	8.57	9.75	8.9	90.1	
P1537-13	13	13	1.18	8.72	9.9	8.53	84.3	
P1537-14	14	14	1.19	8.62	9.81	8.94	89.9	
P1537-15	15	15	1.13	8.68	9.81	8.88	89.3	
P1538-01	20241096-COMPOSITE-35N-N-4-01	16	1.12	8.66	9.78	3.93	32.4	sludge sample
P1538-02	20241097-COMPOSITE-35N-N-4-02	17	1.17	8.57	9.74	4.02	33.3	sludge sample
P1538-03	20241098-COMPOSITE-35N-N-4-03	18	1.12	8.66	9.78	4.17	35.2	sludge sample
P1538-04	20241099-COMPOSITE-35N-N-4-04	19	1.15	8.81	9.96	4.1	33.5	sludge sample
P1538-05	20241100-COMPOSITE-35N-N-4-05	20	1.13	8.40	9.53	4.14	35.8	sludge sample
P1538-06	20241101-COMPOSITE-35N-N-4-06	21	1.15	8.80	9.95	4.23	35.0	sludge sample
P1538-07	P1538-06MS	22	1.15	8.80	9.95	4.23	35.0	sludge sample
P1538-08	P1538-06MSD	23	1.15	8.80	9.95	4.23	35.0	sludge sample
P1538-09	20241102-AMENDED-COMPOSITE-35N-N-4-01	24	1.18	8.56	9.74	4.56	39.5	
P1538-10	20241103-AMENDED-COMPOSITE-35N-N-4-02	25	1.16	8.61	9.77	4.56	39.5	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/22/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 02/21/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 02/22/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB129607

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1538-11	20241104-AMENDED-COMPO SITE-35N-N-4-03	26	1.16	8.40	9.56	4.8	43.3	
P1538-12	20241105-AMENDED-COMPO SITE-35N-N-4-04	27	1.18	8.79	9.97	4.78	41.0	
P1538-13	20241106-AMENDED-COMPO SITE-35N-N-4-05	28	1.13	8.80	9.93	5.00	44.0	
P1538-14	20241107-AMENDED-COMPO SITE-35N-N-4-06	29	1.14	8.85	9.99	5.2	45.9	
P1538-15	P1538-14MS	30	1.14	8.85	9.99	5.2	45.9	
P1538-16	P1538-14MSD	31	1.14	8.85	9.99	5.2	45.9	
P1541-01	40289	32	1.00	1.00	2.00	2.00	100.0	wipe sample
P1541-02	40290	33	1.00	1.00	2.00	2.00	100.0	wipe sample
P1541-03	40291	34	1.00	1.00	2.00	2.00	100.0	wipe sample
P1541-04	40292	35	1.00	1.00	2.00	2.00	100.0	wipe sample
P1542-01	W029020-TH198C-1-1	36	1.00	1.00	2.00	2.00	100.0	oilc
P1542-02	W029020-TH198C-1-2	37	1.00	1.00	2.00	2.00	100.0	oilc
P1543-01	COMP-1	38	1.00	1.00	2.00	2.00	100.0	debris
P1543-02	84-OIL	39	1.00	1.00	2.00	2.00	100.0	oil sample
P1543-03	101	40	1.00	1.00	2.00	2.00	100.0	debris
P1543-04	85-OIL	41	1.00	1.00	2.00	2.00	100.0	oil sample
P1544-01	VNJ218	42	1.15	8.81	9.96	9.03	89.4	
P1544-02	VNJ218	43	1.16	8.78	9.94	9.07	90.1	
P1545-01	287	44	1.16	8.82	9.98	9.45	94.0	
P1545-02	287	45	1.16	8.78	9.94	8.72	86.1	
P1546-01	1313	46	1.00	1.00	2.00	2.00	100.0	oil sample
P1547-01	OR-03-022024	47	1.18	8.76	9.94	9.12	90.6	
P1547-02	OR-03-022024-E2	48	1.14	8.45	9.59	8.78	90.4	
P1549-01	HT3727	49	1.13	8.76	9.89	9.03	90.2	
P1549-03	336	50	1.00	1.00	2.00	2.00	100.0	debris
P1555-01	Guard shack Exterior on poured concrete slab	54	1.00	1.00	2.00	2.00	100.0	caluk



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/22/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 02/21/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 02/22/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB129607

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1555-02	Guard shack Exterlor	55	1.00	1.00	2.00	2.00	100.0	caluk
P1555-03	Guard shack Pipe Pen	56	1.00	1.00	2.00	2.00	100.0	caluk
P1555-04	Guard shack Exterlor	57	1.00	1.00	2.00	2.00	100.0	caluk
P1555-05	Guard shack Exterlor Roof Flashing	58	1.00	1.00	2.00	2.00	100.0	caluk
P1560-01	RBR20010	51	1.12	8.83	9.95	9.77	98.0	
P1560-02	RBR20010-E2	52	1.12	8.83	9.95	9.77	98.0	
P1560-03	CONCRETE-SLAB	53	1.00	1.00	2.00	2.00	100.0	CONCRETE-SLAB

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$