



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129610

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/21/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/21/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	02/21/2024	08:15
4	ICV	1	Water	NA	NA	20.3	7.00	02/21/2024	08:20
5	CCV1	1	Water	NA	NA	20.3	2.01	02/21/2024	08:22
6	P1540-01	1	Water	NA	NA	22.6	4.94	02/21/2024	08:25
7	P1540-01DUP	1	Water	NA	NA	22.7	4.95	02/21/2024	08:30
8	P1540-02	1	Water	NA	NA	22.9	5.29	02/21/2024	08:35
9	P1540-03	1	Water	NA	NA	22.7	5.35	02/21/2024	08:40
10	P1540-04	1	Water	NA	NA	22.7	5.75	02/21/2024	08:45
11	P1540-05	1	Water	NA	NA	22.7	5.00	02/21/2024	08:50
12	P1546-04	1	Water	NA	NA	22.7	6.65	02/21/2024	08:52
13	CCV2	1	Water	NA	NA	20.3	12.02	02/21/2024	08:55

VB 129610

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p14549 WorkList ID : 178032 Department : Wet-Chemistry Date : 02-21-2024 07:49:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1540-01	AUD-1323	Water	pH	Cool 4 deg C	PSEG03	E21	02/20/2024	9040C
P1540-02	AUD-1324	Water	pH	Cool 4 deg C	PSEG03	E21	02/20/2024	9040C
P1540-03	AUD-1325	Water	pH	Cool 4 deg C	PSEG03	E21	02/20/2024	9040C
P1540-04	AUD-1326	Water	pH	Cool 4 deg C	PSEG03	E21	02/20/2024	9040C
P1540-05	AUD-1327	Water	pH	Cool 4 deg C	PSEG03	E21	02/20/2024	9040C
P1546-04	COMP-1	Water	pH	Cool 4 deg C	PSEG03	E21	02/20/2024	9040C

Date/Time 02/21/24 07:50

Raw Sample Received by: SP/WC

Raw Sample Relinquished by: CP Sim

Date/Time 02/21/24

Raw Sample Received by: CP Sim

Raw Sample Relinquished by: SP/WC