



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB129611

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/21/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/21/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	02/21/2024	08:15
4	ICV	1	Water	NA	NA	20.3	7.00	02/21/2024	08:20
5	CCV1	1	Water	NA	NA	20.2	2.02	02/21/2024	09:00
6	P1543-05	1	Water	NA	NA	23.1	7.69	02/21/2024	09:05
7	P1543-06	1	Water	NA	NA	24.4	7.07	02/21/2024	09:10
8	P1543-06DUP	1	Water	NA	NA	24.5	7.08	02/21/2024	09:11
9	CCV2	1	Water	NA	NA	20.3	12.02	02/21/2024	09:12

VB 129611

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1543 water WorkList ID : 178036 Department : Wet-Chemistry Date : 02-21-2024 07:53:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1543-05	85-WATER	Water	Corrosivity	Cool 4 deg C	PSEG03	E21	02/20/2024	9040C
P1543-06	COMP-2	Water	Corrosivity	Cool 4 deg C	PSEG03	E21	02/20/2024	9040C

Date/Time 02-21-24 08:00
Raw Sample Received by: JH (COC)
Raw Sample Relinquished by: CR SM

Date/Time 02-21-24 B1:00
Raw Sample Received by: CR SM
Raw Sample Relinquished by: SD (COC)