



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129612

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/21/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/21/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	02/21/2024	08:15
4	ICV	1	Water	NA	NA	20.3	7.00	02/21/2024	08:22
5	CCV1	1	Water	NA	NA	20.2	2.01	02/21/2024	09:20
6	P1318-04	1	Solid	20.02	20	23.2	7.83	02/21/2024	09:25
7	P1318-04DUP	1	Solid	20.03	20	23.3	7.84	02/21/2024	09:26
8	P1318-05	1	Solid	20.01	20	23.0	7.48	02/21/2024	09:30
9	P1530-03	1	Solid	20.02	20	20.7	7.33	02/21/2024	09:35
10	P1543-01	1	Solid	20.02	20	22.3	7.90	02/21/2024	09:38
11	P1543-02	1	Solid	20.01	20	20.2	4.05	02/21/2024	09:48
12	P1543-03	1	Solid	20.03	20	21.6	7.91	02/21/2024	09:50
13	P1543-04	1	Solid	20.01	20	20.2	4.06	02/21/2024	10:00
14	P1549-02	1	Solid	20.02	20	22.2	7.90	02/21/2024	10:10
15	P1549-04	1	Solid	20.03	20	22.1	7.80	02/21/2024	10:15
16	CCV2	1	Water	NA	NA	20.3	12.02	02/21/2024	10:20

W3129612

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1543

WorkList ID : 178035

Department : Wet-Chemistry

Date : 02-21-2024 07:51:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1318-04	CO-11-GRID-1	Solid	Corrosivity	Cool 4 deg C	WALS01	D11	01/30/2024	9045D
P1318-05	CO-11-GRID-2	Solid	Corrosivity	Cool 4 deg C	WALS01	D11	01/30/2024	9045D
P1530-03	BROAD-GENERAL	Solid	Corrosivity	Cool 4 deg C	TULL02	E11	02/19/2024	9045D
P1543-01	COMP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/20/2024	9045D
P1543-02	84-OIL	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/20/2024	9045D
P1543-03	101	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/20/2024	9045D
P1543-04	85-OIL	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/20/2024	9045D
P1549-02	HT3727	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/20/2024	9045D
P1549-04	336	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/20/2024	9045D

Date/Time

02/21/24 08:00

Raw Sample Received by:

JL WC

Raw Sample Relinquished by:

AL SM

Date/Time

02/21/24

Raw Sample Received by:

JL WC

Raw Sample Relinquished by:

AL SM