

LB 129616


SFAM01.1

INSTRUMENT ID: EV1

Linear

 μ Abs.:

50679



A= 0.0000e+000

B= 1.9291e-004 *slope*C= -1.5472e-002 *y-intercept*

Rho= 0.9989500

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.025	0.025	212	0.000	212					
0.05	0.050					442	0				-8
0.20	0.200	0.185	-0.015	1037	0.0 %	1037					-7
2.50	2.500	2.337	-0.163	12194	0.0 %	12194					3
5.00	5.000	5.125	0.125	26645	0.0 %	26645					4
7.50	7.500	7.767	0.267	40343	0.0 %	40343					-2
10.0	10.000	9.761	-0.239	50679	0.0 %	50679					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	212	-		0 SFAM01.1	PPB	2/21/2024 11:28	Std
	0.2 S01	1037	-		0.2 SFAM01.1	PPB	2/21/2024 11:30	Std
	2.5 S02	12194	-		2.5 SFAM01.1	PPB	2/21/2024 11:32	Std
	5 S03	26645	-		5 SFAM01.1	PPB	2/21/2024 11:37	Std
	7.5 S04	40343	-		7.5 SFAM01.1	PPB	2/21/2024 11:39	Std
	10 S056	50679	-		10 SFAM01.1	PPB	2/21/2024 11:44	Std
ICV018	ICV018	21885	4.2064 -		SFAM01.1	PPB	2/21/2024 11:47	SMPL
ICB018	ICB018	-12	-0.0178 -		SFAM01.1	PPB	2/21/2024 11:50	SMPL
CCV058	CCV058	26566	5.1095 -		SFAM01.1	PPB	2/21/2024 11:54	SMPL
CCB058	CCB058	-28	-0.0209 -		SFAM01.1	PPB	2/21/2024 11:56	SMPL
PB159177BL	PBW177	296	0.0416 -		SFAM01.1	PPB	2/21/2024 11:58	SMPL
P1403-01	MCOAA9	123	0.0083 -		SFAM01.1	PPB	2/21/2024 12:01	SMPL
P1403-02	MCOAC9	122	0.0081 -		SFAM01.1	PPB	2/21/2024 12:03	SMPL
P1403-03	MCOAD9	214	0.0258 -		SFAM01.1	PPB	2/21/2024 12:05	SMPL
P1403-04	MCOAF3	120	0.0077 -		SFAM01.1	PPB	2/21/2024 12:07	SMPL
P1403-05	MCOAB0	199	0.0229 -		SFAM01.1	PPB	2/21/2024 12:10	SMPL
P1403-06	MCOAC3	165	0.0164 -		SFAM01.1	PPB	2/21/2024 12:12	SMPL
P1403-07	MCOAD3	265	0.0356 -		SFAM01.1	PPB	2/21/2024 12:14	SMPL
P1403-08	MCOAE0	146	0.0127 -		SFAM01.1	PPB	2/21/2024 12:17	SMPL
P1403-09	MCOAC1	464	0.074 -		SFAM01.1	PPB	2/21/2024 12:19	SMPL
P1403-10	MCOAC8	267	0.036 -		SFAM01.1	PPB	2/21/2024 12:21	SMPL
P1403-11	MCOAD1	457	0.0727 -		SFAM01.1	PPB	2/21/2024 12:23	SMPL
P1403-12	MCOAD1D	484	0.0779 -		SFAM01.1	PPB	2/21/2024 12:26	SMPL
P1403-13	MCOAD1S	4944	0.9383 -		SFAM01.1	PPB	2/21/2024 12:30	SMPL
P1403-14	MCOAD4	2301	0.4284 -		SFAM01.1	PPB	2/21/2024 12:33	SMPL
P1403-15	MCOAA8	246	0.032 -		SFAM01.1	PPB	2/21/2024 12:35	SMPL
P1403-16	MCOAB1	334	0.049 -		SFAM01.1	PPB	2/21/2024 12:37	SMPL
P1403-17	MCOAC5	344	0.0509 -		SFAM01.1	PPB	2/21/2024 12:39	SMPL
P1403-18	MCOAD7	461	0.0735 -		SFAM01.1	PPB	2/21/2024 12:42	SMPL
P1403-19	MCOAE4	446	0.0706 -		SFAM01.1	PPB	2/21/2024 12:44	SMPL
CCV059	CCV059	25125	4.8315 -		SFAM01.1	PPB	2/21/2024 12:46	SMPL
CCB059	CCB059	-43	-0.0238 -		SFAM01.1	PPB	2/21/2024 12:49	SMPL
P1403-20	MCOAE9	113	0.0063 -		SFAM01.1	PPB	2/21/2024 12:51	SMPL
CCV060	CCV060	25196	4.8452 -		SFAM01.1	PPB	2/21/2024 12:53	SMPL
CCB060	CCB060	-92	-0.0332 -		SFAM01.1	PPB	2/21/2024 12:55	SMPL