

LB129620


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

57118



A= 0.0000e+000

B= 1.7356e-004

C= 5.8056e-002

Rho= 0.9996915

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0)
0.0	0.000	0.072	0.072	78	0.000	78					1
0.2	0.200	0.206	0.006	853	0.0 %	853					3
2.5	2.500	2.452	-0.048	13795	0.0 %	13795					2
5.0	5.000	4.854	-0.146	27635	0.0 %	27635					3
7.5	7.500	7.645	0.145	43712	0.0 %	43712					2
10.0	10.000	9.971	-0.029	57118	0.0 %	57118					0

LB129620 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Feb 2024 13:29:35

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	78	-	0.0007471B	PPB		21 Feb 2024 13:39:23	S	Std
0.2 - 1	50.2	853	-	0.20007471B	PPB		21 Feb 2024 13:41:40	S	Std
2.5 - 1	52.2	13795	-	2.50007471B	PPB		21 Feb 2024 13:43:58	S	Std
5.0 - 1	55	27635	-	5.00007471B	PPB		21 Feb 2024 13:46:15	S	Std
7.5 - 1	55.5	43712	-	7.50007471B	PPB		21 Feb 2024 13:48:34	S	Std
10.0 - 1	57.1	57118	-	10.00007471B	PPB		21 Feb 2024 13:53:49	S	Std
ICV19 - 1	ICV19	22822	4.0190	-7471B	PPB		21 Feb 2024 13:56:42	U	SMPL
ICB19 - 1	ICB19	-162	0.0298	-7471B	PPB		21 Feb 2024 13:58:59	U	SMPL
CCV61 - 1	CCV61	28152	4.9440	-7471B	PPB		21 Feb 2024 14:03:58	U	SMPL
CCB61 - 1	CCB61	-172	0.0282	-7471B	PPB		21 Feb 2024 14:06:15	U	SMPL
CRA - 1	CRA	823	0.2009	-7471B	PPB		21 Feb 2024 14:08:40	U	SMPL
HighStd - 1	HighStd	58921	10.2842	-7471B	PPB		21 Feb 2024 14:10:55	U	SMPL
ChkStd - 1	ChkStd	41667	7.2896	-7471B	PPB		21 Feb 2024 14:13:10	U	SMPL
PB159165BL - 1	PBS	-282	0.0091	-7471B	PPB		21 Feb 2024 14:15:42	U	SMPL
PB159165BS - 1	LCSS	23060	4.0603	-7471B	PPB		21 Feb 2024 14:22:21	U	SMPL
P1523-01 - 1	20240763-COMPOSITE-39N-N-3-01	35092	6.1485	-7471B	PPB		21 Feb 2024 14:24:37	U	SMPL
P1523-02 - 1	20240764-COMPOSITE-39N-N-3-02	35020	6.1360	-7471B	PPB		21 Feb 2024 14:27:02	U	SMPL
P1523-03 - 1	20240765-COMPOSITE-39N-N-3-03	33025	5.7897	-7471B	PPB		21 Feb 2024 14:29:29	U	SMPL
P1523-04 - 1	P1523-03MS	52631	9.1925	-7471B	PPB		21 Feb 2024 14:34:22	U	SMPL
P1523-05 - 1	P1523-03MSD	51045	8.9172	-7471B	PPB		21 Feb 2024 14:36:41	U	SMPL
CCV62 - 1	CCV62	27764	4.8767	-7471B	PPB		21 Feb 2024 14:39:12	U	SMPL
CCB62 - 1	CCB62	-192	0.0247	-7471B	PPB		21 Feb 2024 14:41:33	U	SMPL
P1523-06 - 1	20240766-AMENDED-COMP-39N-N-3-01	35067	6.1441	-7471B	PPB		21 Feb 2024 14:43:56	U	SMPL
P1523-07 - 1	20240767-AMENDED-COMP-39N-N-3-02	32582	5.7129	-7471B	PPB		21 Feb 2024 14:46:12	U	SMPL
P1523-08 - 1	20240768-AMENDED-COMP-39N-N-3-03	31095	5.4548	-7471B	PPB		21 Feb 2024 14:48:39	U	SMPL
P1523-08DUP - 1	20240768-AMENDED-COMP-39N-N-3-03DUP	28444	4.9947	-7471B	PPB		21 Feb 2024 14:51:04	U	SMPL
P1523-09 - 1	P1523-08MS	51830	9.0535	-7471B	PPB		21 Feb 2024 14:53:29	U	SMPL
P1523-10 - 1	P1523-08MSD	54515	9.5195	-7471B	PPB		21 Feb 2024 14:55:56	U	SMPL
P1530-01 - 1	BROAD-GENERAL	24875	4.3753	-7471B	PPB		21 Feb 2024 14:58:24	U	SMPL
PB159166BL - 1	PBS	-154	0.0313	-7471B	PPB		21 Feb 2024 15:00:52	U	SMPL
PB159166BS - 1	LCSS	23899	4.2059	-7471B	PPB		21 Feb 2024 15:05:28	U	SMPL
P1537-01 - 1	1	8508	1.5347	-7471B	PPB		21 Feb 2024 15:07:43	U	SMPL
CCV63 - 1	CCV63	29124	5.1127	-7471B	PPB		21 Feb 2024 15:10:05	U	SMPL
CCB63 - 1	CCB63	-175	0.0277	-7471B	PPB		21 Feb 2024 15:12:23	U	SMPL
P1537-01DUP - 1	1DUP	8252	1.4902	-7471B	PPB		21 Feb 2024 15:14:46	U	SMPL
P1537-01MS - 1	1MS	36125	6.3278	-7471B	PPB		21 Feb 2024 15:24:53	U	SMPL
P1537-01MSD - 1	1MSD	36207	6.3420	-7471B	PPB		21 Feb 2024 15:27:09	U	SMPL
P1537-02 - 1	2	4905	0.9093	-7471B	PPB		21 Feb 2024 15:29:38	U	SMPL
P1537-03 - 1	3	5168	0.9550	-7471B	PPB		21 Feb 2024 15:32:06	U	SMPL
P1537-04 - 1	4	1994	0.4041	-7471B	PPB		21 Feb 2024 15:34:26	U	SMPL
P1537-05 - 1	5	2849	0.5525	-7471B	PPB		21 Feb 2024 15:36:45	U	SMPL
P1537-06 - 1	6	4934	0.9144	-7471B	PPB		21 Feb 2024 15:39:02	U	SMPL
P1537-07 - 1	7	2079	0.4189	-7471B	PPB		21 Feb 2024 15:41:21	U	SMPL
P1537-08 - 1	8	4355	0.8139	-7471B	PPB		21 Feb 2024 15:43:39	U	SMPL
CCV64 - 1	CCV64	29293	5.1420	-7471B	PPB		21 Feb 2024 15:45:55	U	SMPL
CCB64 - 1	CCB64	-89	0.0428	-7471B	PPB		21 Feb 2024 15:48:12	U	SMPL
P1537-09 - 1	9	1796	0.3698	-7471B	PPB		21 Feb 2024 15:50:37	U	SMPL
P1537-10 - 1	10	3704	0.7009	-7471B	PPB		21 Feb 2024 15:52:53	U	SMPL
P1537-11 - 1	11	973	0.2269	-7471B	PPB		21 Feb 2024 15:55:08	U	SMPL
P1537-12 - 1	12	1929	0.3928	-7471B	PPB		21 Feb 2024 15:57:25	U	SMPL
P1537-13 - 1	13	5891	1.0805	-7471B	PPB		21 Feb 2024 15:59:41	U	SMPL
P1537-14 - 1	14	1323	0.2877	-7471B	PPB		21 Feb 2024 16:01:57	U	SMPL
P1537-15 - 1	15	1437	0.3075	-7471B	PPB		21 Feb 2024 16:08:08	U	SMPL
PB159167BL - 1	PBS	18	0.0612	-7471B	PPB		21 Feb 2024 16:10:24	U	SMPL
PB159167BS - 1	LCSS	23577	4.1500	-7471B	PPB		21 Feb 2024 16:14:46	U	SMPL
P1318-01 - 1	CO-11-GRID-1	20275	3.5774	-7471B	PPB		21 Feb 2024 16:17:03	U	SMPL
CCV65 - 1	CCV65	28237	4.9588	-7471B	PPB		21 Feb 2024 16:19:28	U	SMPL
CCB65 - 1	CCB65	-97	0.0412	-7471B	PPB		21 Feb 2024 16:21:51	U	SMPL
P1318-02 - 1	CO-11-GRID-2	13882	2.4674	-7471B	PPB		21 Feb 2024 16:24:17	U	SMPL
P1538-01 - 1	20241096-COMPOSITE-35N-N-4-01	53554	9.3527	-7471B	PPB		21 Feb 2024 16:26:33	U	SMPL
P1538-02 - 1	20241097-COMPOSITE-35N-N-4-02	37875	6.6315	-7471B	PPB		21 Feb 2024 16:28:52	U	SMPL
P1538-03 - 1	20241098-COMPOSITE-35N-N-4-03	49502	8.6494	-7471B	PPB		21 Feb 2024 16:31:24	U	SMPL
P1538-04 - 1	20241099-COMPOSITE-35N-N-4-04	34717	6.0834	-7471B	PPB		21 Feb 2024 16:33:55	U	SMPL
P1538-05 - 1	20241100-COMPOSITE-35N-N-4-05	45101	7.8856	-7471B	PPB		21 Feb 2024 16:36:27	U	SMPL
P1538-06 - 1	20241101-COMPOSITE-35N-N-4-06	50647	8.8481	-7471B	PPB		21 Feb 2024 16:38:54	U	SMPL
P1538-06DUP - 1	20241101-COMPOSITE-35N-N-4-06DUP	53766	9.3895	-7471B	PPB		21 Feb 2024 16:44:24	U	SMPL
P1538-07 - 1	P1538-06MS	73087	12.7427	-7471B	PPB		21 Feb 2024 16:51:53	U	SMPL
P1538-08 - 1	P1538-06MSD	70822	12.3496	-7471B	PPB		21 Feb 2024 16:54:15	U	SMPL
CCV66 - 1	CCV66	28319	4.9730	-7471B	PPB		21 Feb 2024 16:56:49	U	SMPL
CCB66 - 1	CCB66	-132	0.0351	-7471B	PPB		21 Feb 2024 16:59:12	U	SMPL
P1538-09 - 1	20241102-AMENDED-COMPOSITE-35N-N-4-01	43879	7.6733	-7471B	PPB		21 Feb 2024 17:01:38	U	SMPL
P1538-10 - 1	20241103-AMENDED-COMPOSITE-35N-N-4-02	37529	6.5714	-7471B	PPB		21 Feb 2024 17:03:54	U	SMPL
P1538-11 - 1	20241104-AMENDED-COMPOSITE-35N-N-4-03	48526	8.4800	-7471B	PPB		21 Feb 2024 17:06:23	U	SMPL
P1538-12 - 1	20241105-AMENDED-COMPOSITE-35N-N-4-04	34238	6.0003	-7471B	PPB		21 Feb 2024 17:08:53	U	SMPL
P1538-13 - 1	20241106-AMENDED-COMPOSITE-35N-N-4-05	44557	7.7912	-7471B	PPB		21 Feb 2024 17:11:23	U	SMPL
P1538-14 - 1	20241107-AMENDED-COMPOSITE-35N-N-4-06	42768	7.4807	-7471B	PPB		21 Feb 2024 17:13:53	U	SMPL
P1538-15 - 1	P1538-14MS	65199	11.3737	-7471B	PPB		21 Feb 2024 17:16:24	U	SMPL
P1538-16 - 1	P1538-14MSD	64436	11.2413	-7471B	PPB		21 Feb 2024 17:18:54	U	SMPL
P1544-01 - 1	VNJ218	2369	0.4692	-7471B	PPB		21 Feb 2024 17:21:26	U	SMPL
P1545-01 - 1	287	1332	0.2892	-7471B	PPB		21 Feb 2024 17:23:58	U	SMPL
CCV67 - 1	CCV67	28360	4.9801	-7471B	PPB		21 Feb 2024 17:26:16	U	SMPL
CCB67 - 1	CCB67	-88	0.0429	-7471B	PPB		21 Feb 2024 17:28:32	U	SMPL
P1547-01 - 1	OR-03-022024	1966	0.3993	-7471B	PPB		21 Feb 2024 17:30:59	U	SMPL
P1549-01 - 1	HT3727	14086	2.5031	-7471B	PPB		21 Feb 2024 17:33:16	U	SMPL
P1549-03 - 1	336	-62	0.0473	-7471B	PPB		21 Feb 2024 17:35:32	U	SMPL
P1523-08LX5 - 1		8158	1.4739	-7471B	PPB		21 Feb 2024 17:37:52	U	SMPL
P1523-08A - 1		51878	9.0618	-7471B	PPB		21 Feb 2024 17:40:08	U	SMPL
P1537-01LX5 - 1		1758	0.3633	-7471B	PPB		21 Feb 2024 17:42:26	U	SMPL
P1537-01A - 1		35863	6.2823	-7471B	PPB		21 Feb 2024 17:44:57	U	SMPL
P1538-06LX5 - 1		10463	1.8740	-7471B	PPB		21 Feb 2024 17:47:14	U	SMPL
P1538-06A - 1		91681	15.9698	-7471B	PPB		21 Feb 2024 17:49:43	U	SMPL
CCV68 - 1	CCV68	28296	4.9690	-7471B	PPB		21 Feb 2024 17:52:02	U	SMPL
CCB68 - 1	CCB68	-180	0.0268	-7471B	PPB		21 Feb 2024 17:54:38	U	SMPL