



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB129621
Thermometer ID: Flashpoint

Reviewed By: Nikita
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 771.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	02/21/2024	12:55
2	P1540-01		1	13	100.00	>212.0	>212.0	02/21/2024	13:25
3	P1540-01DUP		1	12	100.00	>212.0	>212.0	02/21/2024	13:55
4	P1540-02		1	14	100.00	>212.0	>212.0	02/21/2024	14:25
5	P1540-03		1	12	100.00	>212.0	>212.0	02/21/2024	14:55
6	P1540-04		1	13	100.00	>212.0	>212.0	02/21/2024	15:25
7	P1540-05		1	14	100.00	>212.0	>212.0	02/21/2024	15:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129621

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 02/21/23 flash point WorkList ID : 178065 Department : Wet-Chemistry Date : 02-21-2024 09:44:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1540-01	AUD-1323	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B
P1540-02	AUD-1324	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B
P1540-03	AUD-1325	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B
P1540-04	AUD-1326	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B
P1540-05	AUD-1327	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B

Date/Time 02/21/2024 12:45
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.

Date/Time 02/21/2024 16:25
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.