

LB129627


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46737



A= 0.0000e+000

B= 2.1374e-004

C= 2.2182e-002

Rho= 0.9993513

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.052	0.052	138	0.000	138	0				-
0.2	0.200	0.175	-0.025	717	0.0 %	717					-13
2.5	2.500	2.362	-0.138	10945	0.0 %	10945					-6
5.0	5.000	5.248	0.248	24448	0.0 %	24448					5
7.5	7.500	7.352	-0.148	34291	0.0 %	34291					-2
10.0	10.000	10.012	0.012	46737	0.0 %	46737					0

LB129627 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Feb 2024 11:30:55

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	501	138	-	0.0000	7470A	PPB	22 Feb 2024 12:24:37	S	Std
0.2 - 1	502	717	-	0.2000	7470A	PPB	22 Feb 2024 12:26:54	S	Std
2.5 - 1	52.5	10945	-	2.5000	7470A	PPB	22 Feb 2024 12:29:11	S	Std
5.0 - 1	55	24448	-	5.0000	7470A	PPB	22 Feb 2024 12:36:40	S	Std
7.5 - 1	57.5	34291	-	7.5000	7470A	PPB	22 Feb 2024 12:49:36	S	Std
10.0 - 1	57.5	46737	-	10.0000	7470A	PPB	22 Feb 2024 12:57:11	S	Std
ICV20 - 1	ICV20	19286	4.1444	-	7470A	PPB	22 Feb 2024 13:02:36	U	SMPL
ICB20 - 1	ICB20	79	0.0391	-	7470A	PPB	22 Feb 2024 13:04:52	U	SMPL
CCV69 - 1	CCV69	23069	4.9530	-	7470A	PPB	22 Feb 2024 13:09:44	U	SMPL
CCB69 - 1	CCB69	73	0.0378	-	7470A	PPB	22 Feb 2024 13:12:00	U	SMPL
CRA - 1	CRA	645	0.1600	-	7470A	PPB	22 Feb 2024 13:14:18	U	SMPL
HighStd - 1	HighStd	46059	9.8670	-	7470A	PPB	22 Feb 2024 13:16:33	U	SMPL
ChkStd - 1	ChkStd	31633	6.7835	-	7470A	PPB	22 Feb 2024 13:18:48	U	SMPL
PB159180BL - 1	PBW	-349	-0.0524	-	7470A	PPB	22 Feb 2024 13:21:07	U	SMPL
PB159180BS - 1	LCSW	17683	3.8018	-	7470A	PPB	22 Feb 2024 13:23:27	U	SMPL
P1524-06 - 1	20240766-AMENDED-COMP-39N-N-3-01	1516	0.3462	-	7470A	PPB	22 Feb 2024 13:25:43	U	SMPL
P1524-07 - 1	20240767-AMENDED-COMP-39N-N-3-02	1766	0.3997	-	7470A	PPB	22 Feb 2024 13:28:02	U	SMPL
P1524-08 - 1	20240768-AMENDED-COMP-39N-N-3-03	1208	0.2804	-	7470A	PPB	22 Feb 2024 13:30:19	U	SMPL
P1524-08DUP - 1	20240768-AMENDED-COMP-39N-N-3-03DUP	1304	0.3009	-	7470A	PPB	22 Feb 2024 13:32:37	U	SMPL
P1524-09 - 1	P1524-08MS	18263	3.9258	-	7470A	PPB	22 Feb 2024 13:34:52	U	SMPL
CCV70 - 1	CCV70	22410	4.8122	-	7470A	PPB	22 Feb 2024 13:37:07	U	SMPL
CCB70 - 1	CCB70	59	0.0348	-	7470A	PPB	22 Feb 2024 13:39:25	U	SMPL
P1524-10 - 1	P1524-08MSD	20147	4.3285	-	7470A	PPB	22 Feb 2024 13:41:43	U	SMPL
PB159181BL - 1	PBW	-130	-0.0056	-	7470A	PPB	22 Feb 2024 13:43:58	U	SMPL
PB159181BS - 1	LCSW	17146	3.6870	-	7470A	PPB	22 Feb 2024 13:46:16	U	SMPL
P1529-02 - 1	WA-2	-212	-0.0231	-	7470A	PPB	22 Feb 2024 13:48:32	U	SMPL
P1529-04 - 1	WA-3	-16	0.0188	-	7470A	PPB	22 Feb 2024 13:50:50	U	SMPL
P1529-06 - 1	WA-5	38	0.0303	-	7470A	PPB	22 Feb 2024 13:53:06	U	SMPL
P1529-08 - 1	WA-7	84	0.0401	-	7470A	PPB	22 Feb 2024 13:55:23	U	SMPL
P1529-10 - 1	WA-1	16	0.0256	-	7470A	PPB	22 Feb 2024 13:57:40	U	SMPL
P1531-02 - 1	EDGEMERE	29	0.0284	-	7470A	PPB	22 Feb 2024 13:59:56	U	SMPL
P1531-02DUP - 1	EDGEMEREDUP	33	0.0292	-	7470A	PPB	22 Feb 2024 14:02:12	U	SMPL
CCV71 - 1	CCV71	22674	4.8686	-	7470A	PPB	22 Feb 2024 14:04:28	U	SMPL
CCB71 - 1	CCB71	113	0.0463	-	7470A	PPB	22 Feb 2024 14:06:43	U	SMPL
P1531-02MS - 1	EDGEMEREMS	19110	4.1068	-	7470A	PPB	22 Feb 2024 14:09:01	U	SMPL
P1531-02MSD - 1	EDGEMEREMSD	18387	3.9523	-	7470A	PPB	22 Feb 2024 14:11:17	U	SMPL
PB159192BL - 1	PBW	-270	-0.0355	-	7470A	PPB	22 Feb 2024 14:13:35	U	SMPL
PB159192BS - 1	LCSW	15951	3.4316	-	7470A	PPB	22 Feb 2024 14:15:54	U	SMPL
P1530-03 - 1	BROAD-GENERAL	-183	-0.0169	-	7470A	PPB	22 Feb 2024 14:18:10	U	SMPL
P1543-01 - 1	COMP-1	1	0.0224	-	7470A	PPB	22 Feb 2024 14:20:29	U	SMPL
P1543-02 - 1	84-OIL	183	0.0613	-	7470A	PPB	22 Feb 2024 14:22:46	U	SMPL
P1543-03 - 1	101	56	0.0342	-	7470A	PPB	22 Feb 2024 14:25:03	U	SMPL
P1543-04 - 1	85-OIL	159	0.0562	-	7470A	PPB	22 Feb 2024 14:27:20	U	SMPL
P1549-02 - 1	HT3727	302	0.0867	-	7470A	PPB	22 Feb 2024 14:29:36	U	SMPL
CCV72 - 1	CCV72	21407	4.5978	-	7470A	PPB	22 Feb 2024 14:31:51	U	SMPL
CCB72 - 1	CCB72	139	0.0519	-	7470A	PPB	22 Feb 2024 14:34:07	U	SMPL
P1549-04 - 1	336	12	0.0247	-	7470A	PPB	22 Feb 2024 14:36:25	U	SMPL
P1549-04DUP - 1	336DUP	1	0.0224	-	7470A	PPB	22 Feb 2024 14:38:40	U	SMPL
P1549-04MS - 1	336MS	15611	3.3589	-	7470A	PPB	22 Feb 2024 14:40:56	U	SMPL
P1549-04MSD - 1	336MSD	15483	3.3316	-	7470A	PPB	22 Feb 2024 14:46:16	U	SMPL
PB159193BL - 1	PBW	-183	-0.0169	-	7470A	PPB	22 Feb 2024 14:48:31	U	SMPL
PB159193BS - 1	LCSW	16644	3.5797	-	7470A	PPB	22 Feb 2024 14:53:35	U	SMPL
P1543-05 - 1	85-WATER	185	0.0617	-	7470A	PPB	22 Feb 2024 14:55:51	U	SMPL
P1543-06 - 1	COMP-2	374	0.1021	-	7470A	PPB	22 Feb 2024 14:58:11	U	SMPL
P1543-06DUP - 1	COMP-2DUP	394	0.1064	-	7470A	PPB	22 Feb 2024 15:00:29	U	SMPL
P1543-06MS - 1	COMP-2MS	15482	3.3313	-	7470A	PPB	22 Feb 2024 15:02:45	U	SMPL
CCV73 - 1	CCV73	22643	4.8620	-	7470A	PPB	22 Feb 2024 15:07:39	U	SMPL
CCB73 - 1	CCB73	63	0.0356	-	7470A	PPB	22 Feb 2024 15:09:55	U	SMPL
P1543-06MSD - 1	COMP-2MSD	15240	3.2796	-	7470A	PPB	22 Feb 2024 15:12:14	U	SMPL
PB159202BL - 1	PBW	199	0.0647	-	7470A	PPB	22 Feb 2024 15:14:29	U	SMPL
PB159202BS - 1	LCSW	16603	3.5710	-	7470A	PPB	22 Feb 2024 15:16:48	U	SMPL
P1540-01 - 1	AUD-1323	708	0.1735	-	7470A	PPB	22 Feb 2024 15:34:21	U	SMPL
P1540-01DUP - 1	AUD-1323DUP	626	0.1560	-	7470A	PPB	22 Feb 2024 15:36:37	U	SMPL
P1540-01MS - 1	AUD-1323MS	13934	3.0005	-	7470A	PPB	22 Feb 2024 15:41:43	U	SMPL
P1540-01MSD - 1	AUD-1323MSD	13897	2.9926	-	7470A	PPB	22 Feb 2024 15:43:59	U	SMPL
P1540-02 - 1	AUD-1324	540	0.1376	-	7470A	PPB	22 Feb 2024 15:46:18	U	SMPL
P1540-03 - 1	AUD-1325	2741	0.6081	-	7470A	PPB	22 Feb 2024 15:48:38	U	SMPL
P1540-04 - 1	AUD-1326	1477	0.3379	-	7470A	PPB	22 Feb 2024 15:50:54	U	SMPL
CCV74 - 1	CCV74	21092	4.5304	-	7470A	PPB	22 Feb 2024 15:53:12	U	SMPL
CCB74 - 1	CCB74	210	0.0671	-	7470A	PPB	22 Feb 2024 15:55:27	U	SMPL
P1546-04 - 1	COMP-1	11708	2.5247	-	7470A	PPB	22 Feb 2024 15:57:46	U	SMPL
P1524-08LX5 - 1		179	0.0604	-	7470A	PPB	22 Feb 2024 16:00:02	U	SMPL
P1524-08A - 1		17040	3.6644	-	7470A	PPB	22 Feb 2024 16:02:21	U	SMPL
P1531-02LX5 - 1		-156	-0.0112	-	7470A	PPB	22 Feb 2024 16:04:37	U	SMPL
P1531-02A - 1		17135	3.6847	-	7470A	PPB	22 Feb 2024 16:06:56	U	SMPL
P1549-04LX5 - 1		-187	-0.0178	-	7470A	PPB	22 Feb 2024 16:09:12	U	SMPL
P1549-04A - 1		15485	3.3320	-	7470A	PPB	22 Feb 2024 16:11:31	U	SMPL
P1543-06LX5 - 1		-106	-0.0005	-	7470A	PPB	22 Feb 2024 16:13:48	U	SMPL
P1543-06A - 1		15535	3.3427	-	7470A	PPB	22 Feb 2024 16:16:08	U	SMPL
CCV75 - 1	CCV75	21505	4.6187	-	7470A	PPB	22 Feb 2024 16:18:24	U	SMPL
CCB75 - 1	CCB75	271	0.0801	-	7470A	PPB	22 Feb 2024 16:20:43	U	SMPL
P1540-01LX5 - 1		35	0.0297	-	7470A	PPB	22 Feb 2024 16:25:38	U	SMPL
P1540-01A - 1		16624	3.5754	-	7470A	PPB	22 Feb 2024 16:31:16	U	SMPL
PB159118TB - 1	PB159118TB	-47	0.0121	-	7470A	PPB	22 Feb 2024 16:33:32	U	SMPL
PB159143TB - 1	PB159143TB	139	0.0519	-	7470A	PPB	22 Feb 2024 16:35:51	U	SMPL
PB159160TB - 1	PB159160TB	66	0.0363	-	7470A	PPB	22 Feb 2024 16:38:07	U	SMPL
PB159162TB - 1	PB159162TB	362	0.0996	-	7470A	PPB	22 Feb 2024 16:40:23	U	SMPL
CCV76 - 1	CCV76	22233	4.7743	-	7470A	PPB	22 Feb 2024 16:45:17	U	SMPL
CCB76 - 1	CCB76	114	0.0465	-	7470A	PPB	22 Feb 2024 16:47:33	U	SMPL