



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129628

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.1	02/22/2024	08:25
2	P1543-05		1	12	100.00	>212.0	>212.0	02/22/2024	08:50
3	P1543-05DUP		1	13	100.00	>212.0	>212.0	02/22/2024	09:20
4	P1543-06		1	14	100.00	>212.0	>212.0	02/22/2024	09:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB.129628

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 02/22/24 flashpoint w

WorkList ID : 178092

Department : Wet-Chemistry

Date : 02-22-2024 08:09:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1543-05	85-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B
P1543-06	COMP-2	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B

Date/Time 02/22/2024 8:15

Raw Sample Received by: N.P.W.C

Raw Sample Relinquished by: [Signature]

Date/Time 02/22/2024 10:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: N.P.W.C