



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129645

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	02/23/2024	12:40
2	P1529-01		1	12	100.00	>212.0	>212.0	02/23/2024	13:10
3	P1529-03		1	14	100.00	>212.0	>212.0	02/23/2024	13:40
4	P1529-05		1	13	100.00	>212.0	>212.0	02/23/2024	14:10
5	P1529-07		1	12	100.00	>212.0	>212.0	02/23/2024	14:40
6	P1529-09		1	14	100.00	>212.0	>212.0	02/23/2024	15:10
7	P1531-01		1	13	100.00	>212.0	>212.0	02/23/2024	15:40
8	P1574-14		1	12	100.00	>212.0	>212.0	02/23/2024	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

VP 118545

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss n1396 WorkList ID : 156569 Department : Wet-Chemistry Date : 02-04-2022 09:02:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/10/2022	Water	N1385-01	TSS	Cool 4 deg C	PHAR02	N12	32220008-3	02/03/2022	SM2540 D
02/10/2022	Water	N1386-01	TSS	Cool 4 deg C	PHAR02	N12	15220012-7	02/03/2022	SM2540 D
02/10/2022	Water	N1396-01	TSS	Cool 4 deg C	PSEG04	M11	DSN002	02/03/2022	SM2540 D
02/10/2022	Water	N1396-03	TSS	Cool 4 deg C	PSEG04	M11	DSN001	02/03/2022	SM2540 D
02/10/2022	Water	N1396-05	TSS	Cool 4 deg C	PSEG04	M11	RWS003	02/03/2022	SM2540 D
02/10/2022	Water	N1396-06	TSS	Cool 4 deg C	PSEG04	M11	RWS003	02/03/2022	SM2540 D
02/10/2022	Water	N1396-07	TSS	Cool 4 deg C	PSEG04	M11	TOWER-1	02/03/2022	SM2540 D
02/10/2022	Water	N1396-08	TSS	Cool 4 deg C	PSEG04	M11	TOWER-2	02/03/2022	SM2540 D
02/09/2022	Water	N1416-01	TSS	Cool 4 deg C	TULL01	N13	001-WILLETS-PT-BLVD(FEB)	02/02/2022	SM2540 D
02/09/2022	Water	N1416-02	TSS	Cool 4 deg C	TULL01	N13	002-35TH-AVE(FEB)	02/02/2022	SM2540 D
02/09/2022	Water	N1417-01	TSS	Cool 4 deg C	TULL02	N11	PIT-11-DEWATERING	02/03/2022	SM2540 D

Date/Time 02/10/22 14:00
Raw Sample Received by: JD WCI
Raw Sample Relinquished by: RJ CEST-1066

Date/Time 02/10/22 18:00
Raw Sample Received by: RJ CEST-1066
Raw Sample Relinquished by: JD WCI