



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB129650

Reagent/Standard	Lot/Log #
COD calibration std. 10 ppm	WP105791
COD calibration std. 150 ppm	WP105794
COD calibration std. 50 ppm	WP105792
COD calibration std. 0 ppm	WP105790
COD ICV-LCS std, 50ppm	WP105920
COD calibration std. 100 ppm	WP105793
COD Digestion Vials Low Level 0-150Mg/L	W3067
COD ICV-LCS std, 50ppm	WP106747
COD CCV std, 50ppm	WP106746

Temp In (C): 148	Date In: 02/23/2024	Time In: 09:35
Temp Out (C): 151	Date Out: 02/23/2024	Time Out: 11:35

Intercept: 0.0126

Slope: 1.003

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.013		12/19/2023	14:30
2	CAL2	10	1	Water	11.000	10.955	9.6	12/19/2023	14:30
3	CAL3	50	1	Water	49.000	48.841	-2.3	12/19/2023	14:31
4	CAL4	100	1	Water	100.000	99.688	-0.3	12/19/2023	14:31
5	CAL5	150	1	Water	151.000	150.536	0.4	12/19/2023	14:32



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB129650

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	50.000	49.838	12/19/2023	14:32
2	ICB		NA	NA	1	Water	1.000	0.984	12/19/2023	14:33
3	CCV1	50	NA	NA	1	Water	51.000	50.835	02/23/2024	14:10
4	CCB1		NA	NA	1	Water	0.000	-0.013	02/23/2024	14:10
5	LB129650BL		NA	NA	1	Water	1.000	0.984	02/23/2024	14:11
6	LB129650BS	50	NA	NA	1	Water	51.000	50.835	02/23/2024	14:11
7	P1477-01		NA	NA	10	Water	31.000	30.895	02/23/2024	14:12
8	P1477-01DUP		NA	NA	10	Water	31.000	30.895	02/23/2024	14:12
9	P1477-01MS	50	NA	NA	10	Water	35.000	34.883	02/23/2024	14:13
10	P1477-01MSD	50	NA	NA	10	Water	35.000	34.883	02/23/2024	14:13
11	P1479-01		NA	NA	10	Water	33.000	32.889	02/23/2024	14:14
12	P1562-01		NA	NA	1	Water	131.000	130.596	02/23/2024	14:14
13	P1562-02		NA	NA	10	Water	63.000	62.799	02/23/2024	14:15
14	CCV2	50	NA	NA	1	Water	50.000	49.838	02/23/2024	14:15
15	CCB2		NA	NA	1	Water	1.000	0.984	02/23/2024	14:16