



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129655

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/23/2024	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	02/23/2024	16:01
3	CAL3	1	Water	NA	NA	20.3	10.02	02/23/2024	16:03
4	ICV	1	Water	NA	NA	20.2	7.01	02/23/2024	16:05
5	CCV1	1	Water	NA	NA	20.3	2.01	02/23/2024	16:10
6	P1581-01	1	Water	NA	NA	20.6	4.89	02/23/2024	16:11
7	P1581-01DUP	1	Water	NA	NA	20.7	4.90	02/23/2024	16:12
8	P1583-01	1	Water	NA	NA	24.0	8.46	02/23/2024	16:25
9	P1587-02	1	Water	NA	NA	20.6	9.07	02/23/2024	16:30
10	P1587-04	1	Water	NA	NA	20.8	9.11	02/23/2024	16:35
11	P1587-05	1	Water	NA	NA	20.8	9.06	02/23/2024	16:40
12	CCV2	1	Water	NA	NA	20.3	12.02	02/23/2024	16:45

WB 129655

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1581 WorkList ID : 178140 Department : Wet-Chemistry Date : 02-23-2024 15:34:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1581-01	CELL-A-POOL-WATER	Water	pH	Cool 4 deg C	STER06	E21	02/21/2024	9040C
P1583-01	MANHOLE-COMP	Water	pH	Cool 4 deg C	PSEG03	E21	02/23/2024	9040C
P1587-02	1823-A	Water	pH	Cool 4 deg C	PSEG03	E21	02/23/2024	9040C
P1587-04	1326	Water	pH	Cool 4 deg C	PSEG03	E21	02/23/2024	9040C
P1587-05	1325	Water	pH	Cool 4 deg C	PSEG03	E21	02/23/2024	9040C

Date/Time 02/23/24 15:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 02/23/24 17:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]