



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 02/23/2024 16:45 TEMP1 OUT: 103 °C 02/24/2024 08:00
TEMP2 IN: 104 °C 02/24/2024 08:30 TEMP2 OUT: 103 °C 02/24/2024 10:00
TEMP3 IN: 540 °C 02/24/2024 10:30 TEMP3 OUT: 540 °C 02/24/2024 12:00
TEMP4 IN: 560 °C 02/24/2024 12:30 TEMP4 OUT: 560 °C 02/24/2024 14:00

Run Number: LB129657

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB129657BL	85.6327	85.6327	100	85.6328	85.6328	85.6328	85.6328	0.0000	0
2	P1583-01	90.2458	90.2458	100	90.2944	90.2944	90.2671	90.2671	0.0273	273
3	P1583-01DUP	92.5368	92.5368	100	92.5841	92.5841	92.5555	92.5555	0.0286	286

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105 °C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

VB 129637

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tvs water p1583 WorkList ID : 178139 Department : Wet-Chemistry Date : 02-23-2024 15:34:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1583-01	MANHOLE-COMP	Water	TVS	Cool 4 deg C	PSEG03	E21	02/23/2024	SM2540 G

Date/Time 02/23/24 15:40
Raw Sample Received by: JH (ccc)
Raw Sample Relinquished by: [Signature]

Date/Time 02/23/24 17:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: JH (ccc)