

7470A


LB129669

INSTRUMENT D:CV1

Linear

 μ Abs.:

43175



A= 0.0000e+000

B= 2.3163e-004

C= -4.2272e-002

Rho= 0.9996173

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	-0.016	-0.016	115	0.000	115	0				1
0.2	0.200	0.179	-0.021	955	0.0 %	955					1
2.5	2.500	2.435	-0.065	10693	0.0 %	10693					3
5.0	5.000	5.223	0.223	22730	0.0 %	22730					4
7.5	7.500	7.421	-0.079	32221	0.0 %	32221					1
10.0	10.000	9.958	-0.042	43175	0.0 %	43175					0

LB129669 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Feb 2024 10:28:56

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	115	-	0.0007470A	PPB		26 Feb 2024 10:36:05	S	Std
0.2 - 1	50.2	955	-	0.2007470A	PPB		26 Feb 2024 10:38:23	S	Std
2.5 - 1	52.5	10693	-	2.5007470A	PPB		26 Feb 2024 10:40:40	S	Std
5.0 - 1	55	22730	-	5.0007470A	PPB		26 Feb 2024 10:42:57	S	Std
7.5 - 1	55.5	32221	-	7.5007470A	PPB		26 Feb 2024 10:45:16	S	Std
10.0 - 1	57.5	43175	-	10.0007470A	PPB		26 Feb 2024 10:47:34	S	Std
ICV23 - 1	ICV23	16021	3.6687	-7470A	PPB		26 Feb 2024 10:50:40	U	SMPL
ICB23 - 1	ICB23	-159	-0.0791	-7470A	PPB		26 Feb 2024 10:52:54	U	SMPL
CCV85 - 1	CCV85	22590	5.1903	-7470A	PPB		26 Feb 2024 10:55:12	U	SMPL
CCB85 - 1	CCB85	-208	-0.0905	-7470A	PPB		26 Feb 2024 10:57:28	U	SMPL
CRA - 1	CRA	960	0.1801	-7470A	PPB		26 Feb 2024 11:02:23	U	SMPL
HighStd - 1	HighStd	44284	10.2152	-7470A	PPB		26 Feb 2024 11:04:37	U	SMPL
ChkStd - 1	ChkStd	30407	7.0005	-7470A	PPB		26 Feb 2024 11:06:53	U	SMPL
PB159240BL - 1	PBW	120	-0.0145	-7470A	PPB		26 Feb 2024 11:11:49	U	SMPL
PB159240BS - 1	LCSW	17046	3.9061	-7470A	PPB		26 Feb 2024 11:16:38	U	SMPL
P1539-01 - 1	20241102-AMENDED-COMPOSITE-35N-N-4-01	3392	0.7434	-7470A	PPB		26 Feb 2024 11:18:55	U	SMPL
P1539-02 - 1	20241103-AMENDED-COMPOSITE-35N-N-4-02	1907	0.3994	-7470A	PPB		26 Feb 2024 11:21:13	U	SMPL
P1539-03 - 1	20241104-AMENDED-COMPOSITE-35N-N-4-03	1890	0.3955	-7470A	PPB		26 Feb 2024 11:23:30	U	SMPL
P1539-04 - 1	20241105-AMENDED-COMPOSITE-35N-N-4-04	1882	0.3937	-7470A	PPB		26 Feb 2024 11:25:47	U	SMPL
P1539-05 - 1	20241106-AMENDED-COMPOSITE-35N-N-4-05	2002	0.4215	-7470A	PPB		26 Feb 2024 11:28:02	U	SMPL
CCV86 - 1	CCV86	22288	5.1203	-7470A	PPB		26 Feb 2024 11:30:17	U	SMPL
CCB86 - 1	CCB86	-124	-0.0710	-7470A	PPB		26 Feb 2024 11:32:32	U	SMPL
P1539-06 - 1	20241107-AMENDED-COMPOSITE-35N-N-4-06	1154	0.2250	-7470A	PPB		26 Feb 2024 11:34:51	U	SMPL
P1539-06DUP - 1	20241107-AMENDED-COMPOSITE-35N-N-4-06DUP	1163	0.2271	-7470A	PPB		26 Feb 2024 11:37:06	U	SMPL
P1539-07 - 1	P1539-06MS	16637	3.8114	-7470A	PPB		26 Feb 2024 11:39:22	U	SMPL
P1539-08 - 1	P1539-06MSD	16999	3.8952	-7470A	PPB		26 Feb 2024 11:41:37	U	SMPL
P1564-02 - 1	LKWD-BACKFILL-09	1350	0.2704	-7470A	PPB		26 Feb 2024 11:43:56	U	SMPL
P1564-04 - 1	LKWD-BACKFILL-10	64	-0.0274	-7470A	PPB		26 Feb 2024 11:46:15	U	SMPL
P1564-06 - 1	LKWD-BACKFILL-11	177	-0.0013	-7470A	PPB		26 Feb 2024 11:48:33	U	SMPL
P1564-08 - 1	LKWD-BACKFILL-12	176	-0.0015	-7470A	PPB		26 Feb 2024 11:50:49	U	SMPL
P1564-10 - 1	LKWD-BACKFILL-13	201	0.0043	-7470A	PPB		26 Feb 2024 11:53:06	U	SMPL
P1564-12 - 1	LKWD-BACKFILL-14	109	-0.0170	-7470A	PPB		26 Feb 2024 11:55:22	U	SMPL
CCV87 - 1	CCV87	21749	4.9955	-7470A	PPB		26 Feb 2024 11:57:38	U	SMPL
CCB87 - 1	CCB87	-123	-0.0708	-7470A	PPB		26 Feb 2024 11:59:54	U	SMPL
P1564-14 - 1	LKWD-BACKFILL-15	192	0.0022	-7470A	PPB		26 Feb 2024 12:02:12	U	SMPL
P1564-16 - 1	LKWD-BACKFILL-16	176	-0.0015	-7470A	PPB		26 Feb 2024 12:04:28	U	SMPL
P1564-18 - 1	LKWD-BACKFILL-17	164	-0.0043	-7470A	PPB		26 Feb 2024 12:06:43	U	SMPL
P1564-20 - 1	LKWD-BACKFILL-18	152	-0.0071	-7470A	PPB		26 Feb 2024 12:09:00	U	SMPL
PB159286BL - 1	PBW	264	0.0189	-7470A	PPB		26 Feb 2024 12:11:15	U	SMPL
PB159286BS - 1	LCSW	14988	3.4294	-7470A	PPB		26 Feb 2024 12:13:32	U	SMPL
P1574-01 - 1	SB-1	897	0.1655	-7470A	PPB		26 Feb 2024 12:15:48	U	SMPL
P1574-02 - 1	SB-2	970	0.1824	-7470A	PPB		26 Feb 2024 12:18:06	U	SMPL
P1574-03 - 1	SB-3	712	0.1226	-7470A	PPB		26 Feb 2024 12:20:25	U	SMPL
P1574-04 - 1	SB-4	407	0.0520	-7470A	PPB		26 Feb 2024 12:22:40	U	SMPL
CCV88 - 1	CCV88	19855	4.5567	-7470A	PPB		26 Feb 2024 12:24:56	U	SMPL
CCB88 - 1	CCB88	81	-0.0235	-7470A	PPB		26 Feb 2024 12:27:11	U	SMPL
P1574-05 - 1	SB-5	285	0.0237	-7470A	PPB		26 Feb 2024 12:29:30	U	SMPL
P1574-06 - 1	SB-6	307	0.0286	-7470A	PPB		26 Feb 2024 12:31:46	U	SMPL
P1574-07 - 1	SB-7	325	0.0330	-7470A	PPB		26 Feb 2024 12:34:01	U	SMPL
P1574-08 - 1	SB-8	335	0.0363	-7470A	PPB		26 Feb 2024 12:36:17	U	SMPL
P1574-09 - 1	SB-9	416	0.0541	-7470A	PPB		26 Feb 2024 12:38:33	U	SMPL
P1574-10 - 1	SB-10	477	0.0682	-7470A	PPB		26 Feb 2024 12:40:50	U	SMPL
P1574-11 - 1	SB-11	121	-0.0142	-7470A	PPB		26 Feb 2024 12:43:07	U	SMPL
P1574-11DUP - 1	SB-11DUP	116	-0.0154	-7470A	PPB		26 Feb 2024 12:45:24	U	SMPL
P1574-11MS - 1	SB-11MS	14103	3.2244	-7470A	PPB		26 Feb 2024 12:47:41	U	SMPL
P1574-11MSD - 1	SB-11MSD	16569	3.7956	-7470A	PPB		26 Feb 2024 12:49:58	U	SMPL
CCV89 - 1	CCV89	21013	4.8250	-7470A	PPB		26 Feb 2024 12:52:16	U	SMPL
CCB89 - 1	CCB89	6	-0.0409	-7470A	PPB		26 Feb 2024 12:54:34	U	SMPL
PB159214TB - 1	PB159214TB	68	-0.0268	-7470A	PPB		26 Feb 2024 12:56:53	U	SMPL
PB159216TB - 1	PB159216TB	543	0.0835	-7470A	PPB		26 Feb 2024 12:59:09	U	SMPL
P1539-06LX5 - 1		228	0.0105	-7470A	PPB		26 Feb 2024 13:01:24	U	SMPL
P1539-06A - 1		16398	3.7560	-7470A	PPB		26 Feb 2024 13:03:40	U	SMPL
P1574-11LX5 - 1		-96	-0.0650	-7470A	PPB		26 Feb 2024 13:05:55	U	SMPL
P1574-11A - 1		16407	3.7581	-7470A	PPB		26 Feb 2024 13:08:14	U	SMPL
CCV90 - 1	CCV90	21650	4.9725	-7470A	PPB		26 Feb 2024 13:10:30	U	SMPL
CCB90 - 1	CCB90	-5	-0.0434	-7470A	PPB		26 Feb 2024 13:12:50	U	SMPL