



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129671

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	02/26/2024	08:45
2	P1546-04		1	12	100.00	>212.0	>212.0	02/26/2024	09:15
3	P1546-08		1	11	100.00	>212.0	>212.0	02/26/2024	09:45
4	P1546-09		1	14	100.00	>212.0	>212.0	02/26/2024	10:15
5	P1583-01		1	12	100.00	>212.0	>212.0	02/26/2024	10:45
6	P1588-04		1	11	100.00	>212.0	>212.0	02/26/2024	11:15
7	P1588-05		1	12	100.00	>212.0	>212.0	02/26/2024	11:45
8	P1588-05DUP		1	12	100.00	>212.0	>212.0	02/26/2024	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6129671

WorkList Name : fp-0226

WorkList ID : 178164

Department : Wet-Chemistry

Date : 02-26-2024 08:29:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1546-04	COMP-1	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B
P1546-08	1331	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B
P1546-09	1332	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/20/2024	1010B
P1583-01	MANHOLE-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1588-04	1279	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1588-05	1281	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B

Date/Time 02/26/2024 08:35
 Raw Sample Received by: RM (CWC)
 Raw Sample Relinquished by: JG (CWC)

Date/Time 02/26/2024 12:50
 Raw Sample Received by: JG (CWC)
 Raw Sample Relinquished by: RM (CWC)