



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129674

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/26/2024	13:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/26/2024	13:11
3	CAL3	1	Water	NA	NA	20.3	10.02	02/26/2024	13:12
4	ICV	1	Water	NA	NA	20.4	7.02	02/26/2024	13:15
5	CCV1	1	Water	NA	NA	20.3	2.01	02/26/2024	13:20
6	P1599-01	1	Water	NA	NA	22.6	8.26	02/26/2024	13:25
7	P1599-01DUP	1	Water	NA	NA	22.7	8.27	02/26/2024	13:26
8	CCV2	1	Water	NA	NA	20.3	12.02	02/26/2024	13:30

183 129674

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1599 WorkList ID : 178193 Department : Wet-Chemistry Date : 02-26-2024 12:55:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1599-01	34625	Water	pH	Cool 4 deg C	PSEG03	E21	02/26/2024	9040C

Date/Time 02/26/24 13:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 02/26/24 16:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]