



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129675

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/26/2024	13:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/26/2024	13:11
3	CAL3	1	Water	NA	NA	20.3	10.02	02/26/2024	13:12
4	ICV	1	Water	NA	NA	20.4	7.02	02/26/2024	13:15
5	CCV1	1	Water	NA	NA	20.4	2.02	02/26/2024	13:38
6	P1584-01	1	Solid	20.02	20	22.9	9.04	02/26/2024	13:45
7	P1584-01DUP	1	Solid	20.03	20	22.8	9.05	02/26/2024	13:46
8	CCV2	1	Water	NA	NA	20.3	12.02	02/26/2024	13:50

VB129675

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1584 WorkList ID : 178194 Department : Wet-Chemistry Date : 02-26-2024 12:57:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1584-01	AU-06-02232024	Solid	pH	Cool 4 deg C	PSEG05	E21	02/23/2024	9045D

Date/Time 02/26/24 13:00
Raw Sample Received by: Jd Cudde
Raw Sample Relinquished by: Jd Cudde

Date/Time 02/26/24 16:00
Raw Sample Received by: Jd Cudde
Raw Sample Relinquished by: Jd Cudde