


LB129678

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
 B= 1.9425e-004 *slope*
 C= 5.7902e-002 *y-intercept*
 Rho= 0.9992634
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/d</i>
0.00	0.000	0.083	0.083	128	0.000	128					<i>=</i>
0.05	0.050					442	0				<i>0</i>
0.20	0.200	0.200	0.000	733	0.0 %	733					<i>1</i>
2.50	2.500	2.522	0.022	12684	0.0 %	12684					<i>-1</i>
5.00	5.000	4.958	-0.042	25224	0.0 %	25224					<i>-4</i>
7.50	7.500	7.235	-0.265	36947	0.0 %	36947					<i>2</i>
10.0	10.000	10.202	0.202	52223	0.0 %	52223					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	128	-		0 SFAM01.1	PPB	2/29/2024 11:38	Std
	0.2 S01	733	-		0.2 SFAM01.1	PPB	2/29/2024 11:40	Std
	2.5 S02	12684	-		2.5 SFAM01.1	PPB	2/29/2024 11:42	Std
	5 S03	25224	-		5 SFAM01.1	PPB	2/29/2024 11:45	Std
	7.5 S04	36947	-		7.5 SFAM01.1	PPB	2/29/2024 11:47	Std
	10 S05	52223	-		10 SFAM01.1	PPB	2/29/2024 11:52	Std
ICV024	ICV024	19967	3.9366 -		SFAM01.1	PPB	2/29/2024 11:55	SMPL
ICB024	ICB024	-39	0.0503 -		SFAM01.1	PPB	2/29/2024 11:57	SMPL
CCV091	CCV091	25508	5.0129 -		SFAM01.1	PPB	2/29/2024 11:59	SMPL
CCB091	CCB091	-52	0.0478 -		SFAM01.1	PPB	2/29/2024 12:02	SMPL
PB159300BL	PBW300	97	0.0767 -		SFAM01.1	PPB	2/29/2024 12:04	SMPL
P1381-01	CODQ7	111	0.0795 -		SFAM01.1	PPB	2/29/2024 12:06	SMPL
P1381-02	CODR1	121	0.0814 -		SFAM01.1	PPB	2/29/2024 12:08	SMPL
P1381-03	CODR3	123	0.0818 -		SFAM01.1	PPB	2/29/2024 12:11	SMPL
P1381-04	CODR3D	142	0.0855 -		SFAM01.1	PPB	2/29/2024 12:13	SMPL
P1381-05	CODR3S	6417	1.3044 -		SFAM01.1	PPB	2/29/2024 12:18	SMPL
P1381-06	CODR6	150	0.087 -		SFAM01.1	PPB	2/29/2024 12:20	SMPL
P1381-07	CODR9	150	0.087 -		SFAM01.1	PPB	2/29/2024 12:22	SMPL
P1381-08	CODS2	142	0.0855 -		SFAM01.1	PPB	2/29/2024 12:25	SMPL
P1381-09	CODS5	135	0.0841 -		SFAM01.1	PPB	2/29/2024 12:27	SMPL
P1381-10	CODS8	138	0.0847 -		SFAM01.1	PPB	2/29/2024 12:29	SMPL
P1381-11	CODT1	139	0.0849 -		SFAM01.1	PPB	2/29/2024 12:32	SMPL
P1381-12	CODT4	199	0.0966 -		SFAM01.1	PPB	2/29/2024 12:34	SMPL
P1381-13	CODT8	218	0.1002 -		SFAM01.1	PPB	2/29/2024 12:36	SMPL
P1381-14	CODW1	186	0.094 -		SFAM01.1	PPB	2/29/2024 12:38	SMPL
P1381-15	CODY4	190	0.0948 -		SFAM01.1	PPB	2/29/2024 12:41	SMPL
P1381-16	CODW4	185	0.0938 -		SFAM01.1	PPB	2/29/2024 12:43	SMPL
P1381-17	CODW7	193	0.0954 -		SFAM01.1	PPB	2/29/2024 12:45	SMPL
P1381-18	CODX0	216	0.0999 -		SFAM01.1	PPB	2/29/2024 12:47	SMPL
P1381-19	CODX3	215	0.0997 -		SFAM01.1	PPB	2/29/2024 12:50	SMPL
CCV092	CCV092	25188	4.9508 -		SFAM01.1	PPB	2/29/2024 12:52	SMPL
CCB092	CCB092	-85	0.0414 -		SFAM01.1	PPB	2/29/2024 12:54	SMPL
P1381-20	CODY2	319	0.1199 -		SFAM01.1	PPB	2/29/2024 12:57	SMPL
P1381-21	CODX6	305	0.1171 -		SFAM01.1	PPB	2/29/2024 12:59	SMPL
P1381-22	CODX9	334	0.1228 -		SFAM01.1	PPB	2/29/2024 13:01	SMPL
PB159301BL	PBW301	-14	0.0552 -		SFAM01.1	PPB	2/29/2024 13:06	SMPL
P1382-03	CODR4	162	0.0894 -		SFAM01.1	PPB	2/29/2024 13:13	SMPL
P1382-04	CODR4D	145	0.0861 -		SFAM01.1	PPB	2/29/2024 13:16	SMPL
P1382-05	CODR4S	5009	1.0309 -		SFAM01.1	PPB	2/29/2024 13:18	SMPL
P1382-06	CODR7	95	0.0764 -		SFAM01.1	PPB	2/29/2024 13:20	SMPL
P1382-07	CODS0	167	0.0903 -		SFAM01.1	PPB	2/29/2024 13:22	SMPL
P1382-08	CODS3	215	0.0997 -		SFAM01.1	PPB	2/29/2024 13:25	SMPL
P1382-09	CODS6	218	0.1002 -		SFAM01.1	PPB	2/29/2024 13:27	SMPL
P1382-10	CODS9	221	0.1008 -		SFAM01.1	PPB	2/29/2024 13:29	SMPL
P1382-11	CODT2	10	0.0598 -		SFAM01.1	PPB	2/29/2024 13:31	SMPL
P1382-12	CODT5	187	0.0942 -		SFAM01.1	PPB	2/29/2024 13:34	SMPL

LB129678

INSTRUMENT ID : CV1

P1382-13	CODT9	146	0.0863 -	SFAM01.1	PPB	2/29/2024 13:36 SMPL
P1382-14	CODW2	162	0.0894 -	SFAM01.1	PPB	2/29/2024 13:38 SMPL
P1382-15	CODY5	94	0.0762 -	SFAM01.1	PPB	2/29/2024 13:40 SMPL
P1382-16	CODW5	153	0.0876 -	SFAM01.1	PPB	2/29/2024 13:43 SMPL
CCV093	CCV093	25181	4.9494 -	SFAM01.1	PPB	2/29/2024 13:45 SMPL
CCB093	CCB093	-41	0.0499 -	SFAM01.1	PPB	2/29/2024 13:47 SMPL
P1382-17	CODW8	99	0.0771 -	SFAM01.1	PPB	2/29/2024 13:50 SMPL
P1382-18	CODX1	116	0.0804 -	SFAM01.1	PPB	2/29/2024 13:52 SMPL
P1382-19	CODX4	134	0.0839 -	SFAM01.1	PPB	2/29/2024 13:54 SMPL
P1382-20	CODY3	190	0.0948 -	SFAM01.1	PPB	2/29/2024 13:56 SMPL
P1382-21	CODX7	140	0.0851 -	SFAM01.1	PPB	2/29/2024 14:02 SMPL
P1382-22	CODY0	74	0.0723 -	SFAM01.1	PPB	2/29/2024 14:04 SMPL
PB159302BL	PBW302	30	0.0637 -	SFAM01.1	PPB	2/29/2024 14:06 SMPL
P1418-01	CODZ3	161	0.0892 -	SFAM01.1	PPB	2/29/2024 14:09 SMPL
P1418-02	CODZ0	215	0.0997 -	SFAM01.1	PPB	2/29/2024 14:11 SMPL
P1418-03	CODZ1	282	0.1127 -	SFAM01.1	PPB	2/29/2024 14:13 SMPL
P1418-04	CODZ2	255	0.1074 -	SFAM01.1	PPB	2/29/2024 14:15 SMPL
P1418-05	CODZ2D	212	0.0991 -	SFAM01.1	PPB	2/29/2024 14:18 SMPL
P1418-06	CODZ2S	6490	1.3186 -	SFAM01.1	PPB	2/29/2024 14:20 SMPL
P1382-01	CODQ8	196	0.096 -	SFAM01.1	PPB	2/29/2024 14:25 SMPL
P1382-02	CODR2	137	0.0845 -	SFAM01.1	PPB	2/29/2024 14:27 SMPL
CCV094	CCV094	25126	4.9387 -	SFAM01.1	PPB	2/29/2024 14:30 SMPL
CCB094	CCB094	-41	0.0499 -	SFAM01.1	PPB	2/29/2024 14:32 SMPL