



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129694

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/27/2024	14:00
2	CAL2	1	Water	NA	NA	20.3	7.00	02/27/2024	14:01
3	CAL3	1	Water	NA	NA	20.4	10.02	02/27/2024	14:05
4	ICV	1	Water	NA	NA	20.5	7.02	02/27/2024	14:10
5	CCV1	1	Water	NA	NA	20.3	2.01	02/27/2024	14:15
6	P1585-02	1	Solid	20.02	20	24.7	5.30	02/27/2024	14:20
7	P1585-02DUP	1	Solid	20.03	20	24.8	5.31	02/27/2024	14:21
8	P1602-03	1	Solid	20.04	20	25.4	6.12	02/27/2024	14:30
9	P1602-06	1	Solid	20.02	20	24.6	5.72	02/27/2024	14:35
10	P1602-09	1	Solid	20.01	20	24.1	5.74	02/27/2024	14:40
11	P1602-12	1	Solid	20.02	20	24.2	6.53	02/27/2024	14:45
12	CCV2	1	Water	NA	NA	20.3	12.02	02/27/2024	14:50