



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129700

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	02/27/2024	09:10
2	P1587-02		1	12	100.00	>212.0	>212.0	02/27/2024	09:40
3	P1587-02DUP		1	12	100.00	>212.0	>212.0	02/27/2024	10:10
4	P1587-06		1	14	100.00	>212.0	>212.0	02/27/2024	10:40
5	P1587-09		1	13	100.00	>212.0	>212.0	02/27/2024	11:10
6	P1587-10		1	11	100.00	>212.0	>212.0	02/27/2024	11:40
7	P1587-11		1	14	100.00	>212.0	>212.0	02/27/2024	12:10
8	P1587-12		1	12	100.00	>212.0	>212.0	02/27/2024	12:40
9	P1587-13		1	12	100.00	>212.0	>212.0	02/27/2024	13:10
10	P1587-14		1	14	100.00	>212.0	>212.0	02/27/2024	13:40
11	P1587-15		1	14	100.00	>212.0	>212.0	02/27/2024	14:10
12	P1599-01		1	12	100.00	>212.0	>212.0	02/27/2024	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66129700

WorkList Name : FP*0226

WorkList ID : 178196

Department : Wet-Chemistry

Date : 02-26-2024 15:52:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1587-02	1823-A	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1587-06	1325-1326	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1587-09	1826-A	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1587-10	1653	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1587-11	1824-A	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1587-12	1825-A	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1587-13	122	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1587-14	123	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1587-15	124	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/23/2024	1010B
P1599-01	34625	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/26/2024	1010B

Date/Time 02/27/2024 09:00
 Raw Sample Received by: RM (wv)
 Raw Sample Relinquished by: MCoDC1

Date/Time 02/27/2024 15:10
 Raw Sample Received by: SGL GDC1
 Raw Sample Relinquished by: RM (wv)