



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129707

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/28/2024	11:37
2	CAL2	1	Water	NA	NA	20.3	7.00	02/28/2024	11:39
3	CAL3	1	Water	NA	NA	20.4	10.02	02/28/2024	11:40
4	ICV	1	Water	NA	NA	20.2	7.00	02/28/2024	11:45
5	CCV1	1	Water	NA	NA	20.3	2.01	02/28/2024	11:50
6	P1619-01	1	Solid	20.02	20	24.6	5.07	02/28/2024	12:00
7	P1619-01DUP	1	Solid	20.03	20	24.7	5.08	02/28/2024	12:02
8	P1619-02	1	Solid	20.05	20	24.0	5.45	02/28/2024	12:05
9	P1619-03	1	Solid	20.02	20	25.0	5.52	02/28/2024	12:15
10	P1619-04	1	Solid	20.03	20	24.7	5.14	02/28/2024	12:20
11	P1619-05	1	Solid	20.02	20	23.7	5.32	02/28/2024	12:25
12	CCV2	1	Water	NA	NA	20.3	12.02	02/28/2024	12:30

W3 124707

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s p1619 WorkList ID : 178252 Department : Wet-Chemistry Date : 02-28-2024 11:28:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1619-01	SPILL-1	Solid	pH	Cool 4 deg C	PSEG03	E21	02/28/2024	9045D
P1619-02	SPILL-2	Solid	pH	Cool 4 deg C	PSEG03	E21	02/28/2024	9045D
P1619-03	SPILL-3	Solid	pH	Cool 4 deg C	PSEG03	E21	02/28/2024	9045D
P1619-04	SPILL-4	Solid	pH	Cool 4 deg C	PSEG03	E21	02/28/2024	9045D
P1619-05	SPILL-5	Solid	pH	Cool 4 deg C	PSEG03	E21	02/28/2024	9045D

02/28/24 11:30

Date/Time
Raw Sample Received by: SBW
Raw Sample Relinquished by: SBW

02/28/24 14:00
Date/Time
Raw Sample Received by: SBW
Raw Sample Relinquished by: SBW