



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB129711
Thermometer ID: Flashpoint

Reviewed By: Rubina
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 760.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	02/28/2024	09:00
2	P1613-01		1	14	100.00	>212.0	>212.0	02/28/2024	09:30
3	P1613-01DUP		1	14	100.00	>212.0	>212.0	02/28/2024	10:00
4	P1613-02		1	11	100.00	>212.0	>212.0	02/28/2024	10:30
5	P1613-03		1	13	100.00	>212.0	>212.0	02/28/2024	11:00
6	P1613-04		1	12	100.00	>212.0	>212.0	02/28/2024	12:30
7	P1613-05		1	14	100.00	>212.0	>212.0	02/28/2024	13:00
8	P1613-06		1	12	100.00	>212.0	>212.0	02/28/2024	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16129711

WorkList Name : fp-0228

WorkList ID : 178246

Department : Wet-Chemistry

Date : 02-28-2024 08:24:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1613-01	1847	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/27/2024	1010B
P1613-02	1848	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/27/2024	1010B
P1613-03	1849	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/27/2024	1010B
P1613-04	1850	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/27/2024	1010B
P1613-05	1851	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/27/2024	1010B
P1613-06	609	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/27/2024	1010B

Date/Time 02/28/2024 08:45
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: JH(611)

Date/Time 02/28/2024 13:40
 Raw Sample Received by: JH(611)
 Raw Sample Relinquished by: RM CWL