



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/29/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:15
In Date: 02/28/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:50
Out Date: 02/29/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB129713

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1631-01	MF7A20	1	1.15	8.43	9.58	6.67	65.5	
P1631-02	MF7A20D	2	1.15	8.43	9.58	6.67	65.5	
P1631-03	MF7A20S	3	1.15	8.43	9.58	6.67	65.5	
P1631-04	MF7A21	4	1.19	8.41	9.6	7.75	78.0	
P1631-05	MF7A22	5	1.17	8.57	9.74	6.77	65.3	
P1631-06	MF7A23	6	1.16	8.77	9.93	7.19	68.8	
P1631-07	MF7A24	7	1.16	8.81	9.97	8.06	78.3	
P1631-08	MF7A25	8	1.15	8.70	9.85	7.78	76.2	
P1631-09	MF7A26	9	1.14	8.81	9.95	7.43	71.4	
P1631-10	MF7A27	10	1.18	8.42	9.6	6.91	68.1	
P1631-11	MF7A28	11	1.12	8.80	9.92	7.19	69.0	
P1631-12	MF7A29	12	1.18	8.81	9.99	8.31	80.9	
P1631-13	MF7A30	13	1.18	8.64	9.82	6.88	66.0	
P1631-14	MF7A31	14	1.18	8.64	9.82	7.94	78.2	
P1631-15	MF7A32	15	1.19	8.55	9.74	7.88	78.2	
P1631-16	MF7A33	16	1.12	8.56	9.68	7.87	78.9	
P1631-17	MF7A34	17	1.18	8.64	9.82	8.05	79.5	
P1631-18	MF7A35	18	1.19	8.57	9.76	5.61	51.6	
P1631-19	MF7A36	19	1.18	8.80	9.98	6.00	54.8	
P1631-20	MF7A37	20	1.14	8.40	9.54	8.17	83.7	
P1631-21	MF7A38	21	1.11	8.82	9.93	7.62	73.8	
P1631-22	MF7A39	22	1.18	8.81	9.99	7.05	66.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$