

LB129727


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

62347



A= 0.0000e+000

B= 1.5983e-004

C= 5.5061e-003

Rho= 0.9999553

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/v
0.0	0.000	0.044	0.044	243	0.000	243					
0.2	0.200	0.165	-0.035	995	0.0 %	995					-18
2.5	2.500	2.465	-0.035	15388	0.0 %	15388					-1
5.0	5.000	5.014	0.014	31336	0.0 %	31336					0
7.5	7.500	7.542	0.042	47155	0.0 %	47155					0
10.0	10.000	9.970	-0.030	62347	0.0 %	62347					0

LB1297427 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Feb 2024 15:07:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	243	-	0.000	7471B	PPB	29 Feb 2024 15:33:06	S	Std
0.2 - 1	51.2	995	-	0.200	7471B	PPB	29 Feb 2024 15:35:23	S	Std
2.5 - 1	52.5	15388	-	2.500	7471B	PPB	29 Feb 2024 15:37:40	S	Std
5.0 - 1	55	31336	-	5.000	7471B	PPB	29 Feb 2024 15:39:57	S	Std
7.5 - 1	57.5	47155	-	7.500	7471B	PPB	29 Feb 2024 15:42:17	S	Std
10.0 - 1	514	62347	-	10.000	7471B	PPB	29 Feb 2024 15:50:08	S	Std
ICV25 - 1	ICV25	24855	3.9780	-	7471B	PPB	29 Feb 2024 15:52:54	U	SMPL
ICB25 - 1	ICB25	-117	-0.0132	-	7471B	PPB	29 Feb 2024 15:55:11	U	SMPL
CCV95 - 1	CCV95	31029	4.9648	-	7471B	PPB	29 Feb 2024 15:57:31	U	SMPL
CCB95 - 1	CCB95	-97	-0.0100	-	7471B	PPB	29 Feb 2024 15:59:46	U	SMPL
CRA - 1	CRA	1025	0.1693	-	7471B	PPB	29 Feb 2024 16:02:09	U	SMPL
HighStd - 1	HighStd	62801	10.0428	-	7471B	PPB	29 Feb 2024 16:04:25	U	SMPL
ChkStd - 1	ChkStd	44361	7.0956	-	7471B	PPB	29 Feb 2024 16:06:40	U	SMPL
PB159351BL - 1	PBS	182	0.0346	-	7471B	PPB	29 Feb 2024 16:09:10	U	SMPL
PB159351BS - 1	LCSS	29177	4.6688	-	7471B	PPB	29 Feb 2024 16:11:37	U	SMPL
P1584-01 - 1	AU-06-02232024	8080	1.2969	-	7471B	PPB	29 Feb 2024 16:13:53	U	SMPL
P1585-01 - 1	VNJ-232	1978	0.3216	-	7471B	PPB	29 Feb 2024 16:16:16	U	SMPL
P1585-01DUP - 1	VNJ-232DUP	1797	0.2927	-	7471B	PPB	29 Feb 2024 16:21:31	U	SMPL
P1585-01MS - 1	VNJ-232MS	24909	3.9866	-	7471B	PPB	29 Feb 2024 16:28:38	U	SMPL
P1585-01MSD - 1	VNJ-232MSD	25373	4.0608	-	7471B	PPB	29 Feb 2024 16:30:57	U	SMPL
CCV96 - 1	CCV96	30932	4.9493	-	7471B	PPB	29 Feb 2024 16:33:18	U	SMPL
CCB96 - 1	CCB96	-38	-0.0006	-	7471B	PPB	29 Feb 2024 16:35:39	U	SMPL
P1588-03 - 1	1838-1839	2176	0.3533	-	7471B	PPB	29 Feb 2024 16:38:02	U	SMPL
P1600-01 - 1	SPEEDY-DRY-DRUM	1465	0.2397	-	7471B	PPB	29 Feb 2024 16:40:17	U	SMPL
P1602-01 - 1	WC-1	5149	0.8285	-	7471B	PPB	29 Feb 2024 16:42:33	U	SMPL
P1602-04 - 1	WC-2	3010	0.4866	-	7471B	PPB	29 Feb 2024 16:44:49	U	SMPL
P1602-07 - 1	WC-3	3157	0.5101	-	7471B	PPB	29 Feb 2024 16:47:07	U	SMPL
P1602-10 - 1	WC-4	6932	1.1134	-	7471B	PPB	29 Feb 2024 16:49:24	U	SMPL
P1610-01 - 1	SU-02-02272024	1697	0.2767	-	7471B	PPB	29 Feb 2024 16:51:42	U	SMPL
P1611-01 - 1	TR-05-022724	3945	0.6360	-	7471B	PPB	29 Feb 2024 16:54:02	U	SMPL
P1612-01 - 1	OK-02-022724	994	0.1644	-	7471B	PPB	29 Feb 2024 16:56:18	U	SMPL
CCV97 - 1	CCV97	31591	5.0546	-	7471B	PPB	29 Feb 2024 16:58:35	U	SMPL
CCB97 - 1	CCB97	92	0.0202	-	7471B	PPB	29 Feb 2024 17:00:51	U	SMPL
P1585-01LX5 - 1		378	0.0659	-	7471B	PPB	29 Feb 2024 17:03:11	U	SMPL
P1585-01A - 1		28939	4.6307	-	7471B	PPB	29 Feb 2024 17:05:27	U	SMPL
CCV098 - 1	CCV098	31181	4.9891	-	7471B	PPB	29 Feb 2024 17:07:43	U	SMPL
CCB098 - 1	CCB098	25	0.0095	-	7471B	PPB	29 Feb 2024 17:10:03	U	SMPL