

LB129733


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4705E



A= 0.0000e+000

B= 2.1126e-004

C= 2.0648e-002

Rho= 0.9998673

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.049	0.049	134	0.000	134	0				-
0.2	0.200	0.189	-0.011	796	0.0 %	796					-6
2.5	2.500	2.464	-0.036	11564	0.0 %	11564					-1
5.0	5.000	4.929	-0.071	23235	0.0 %	23235					-1
7.5	7.500	7.607	0.107	35909	0.0 %	35909					-1
10.0	10.000	9.962	-0.038	47058	0.0 %	47058					0

LB129733 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Feb 2024 17:28:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	134	-	0.0000	7470A	PPB	29 Feb 2024 17:29:44	S	Std
0.2 - 1	50.2	796	-	0.2000	7470A	PPB	29 Feb 2024 17:32:01	S	Std
2.5 - 1	52.5	11564	-	2.5000	7470A	PPB	29 Feb 2024 17:34:18	S	Std
5.0 - 1	55	23235	-	5.0000	7470A	PPB	29 Feb 2024 17:36:35	S	Std
7.5 - 1	57.5	35909	-	7.5000	7470A	PPB	29 Feb 2024 17:38:53	S	Std
10.0 - 1	110	47058	-	10.0000	7470A	PPB	29 Feb 2024 17:41:11	S	Std
ICV26 - 1	ICV26	20090	4.2649	-	7470A	PPB	29 Feb 2024 17:51:33	U	SMPL
ICB26 - 1	ICB26	-122	-0.0051	-	7470A	PPB	29 Feb 2024 17:53:47	U	SMPL
CCV01 - 1	CCV01	21299	4.5203	-	7470A	PPB	29 Feb 2024 17:56:03	U	SMPL
CCB01 - 1	CCB01	-155	-0.0121	-	7470A	PPB	29 Feb 2024 17:58:18	U	SMPL
CRA - 1	CRA	695	0.1675	-	7470A	PPB	29 Feb 2024 18:00:36	U	SMPL
HighStd - 1	HighStd	46146	9.7696	-	7470A	PPB	29 Feb 2024 18:02:51	U	SMPL
ChkStd - 1	ChkStd	31038	6.5778	-	7470A	PPB	29 Feb 2024 18:05:06	U	SMPL
PB159319BL - 1	PBW	-297	-0.0421	-	7470A	PPB	29 Feb 2024 18:07:25	U	SMPL
PB159319BS - 1	LCSW	16025	3.4061	-	7470A	PPB	29 Feb 2024 18:09:44	U	SMPL
P1585-02 - 1	VNJ-232	-168	-0.0148	-	7470A	PPB	29 Feb 2024 18:12:00	U	SMPL
P1585-02DUP - 1	VNJ-232DUP	-22	0.0160	-	7470A	PPB	29 Feb 2024 18:14:19	U	SMPL
P1585-02MS - 1	VNJ-232MS	19582	4.1576	-	7470A	PPB	29 Feb 2024 18:16:36	U	SMPL
P1585-02MSD - 1	VNJ-232MSD	18116	3.8479	-	7470A	PPB	29 Feb 2024 18:18:53	U	SMPL
PB159337BL - 1	PBW	-189	-0.0193	-	7470A	PPB	29 Feb 2024 18:21:10	U	SMPL
CCV02 - 1	CCV02	22539	4.7823	-	7470A	PPB	29 Feb 2024 18:23:27	U	SMPL
CCB02 - 1	CCB02	-165	-0.0142	-	7470A	PPB	29 Feb 2024 18:25:42	U	SMPL
PB159337BS - 1	LCSW	17674	3.7545	-	7470A	PPB	29 Feb 2024 18:28:00	U	SMPL
P1583-01 - 1	MANHOLE-COMP	131	0.0483	-	7470A	PPB	29 Feb 2024 18:30:15	U	SMPL
P1587-02 - 1	1823-A	347	0.0940	-	7470A	PPB	29 Feb 2024 18:32:32	U	SMPL
P1587-06 - 1	1325-1326	153	0.0530	-	7470A	PPB	29 Feb 2024 18:34:48	U	SMPL
P1587-08 - 1	1255-1259	50	0.0312	-	7470A	PPB	29 Feb 2024 18:37:04	U	SMPL
P1587-09 - 1	1826-A	862	0.2028	-	7470A	PPB	29 Feb 2024 18:39:20	U	SMPL
P1587-09DUP - 1	1826-ADUP	862	0.2028	-	7470A	PPB	29 Feb 2024 18:49:21	U	SMPL
P1587-09MS - 1	1826-AMS	19357	4.1101	-	7470A	PPB	29 Feb 2024 18:54:09	U	SMPL
P1587-09MSD - 1	1826-AMSD	21573	4.5782	-	7470A	PPB	29 Feb 2024 18:56:26	U	SMPL
P1588-05 - 1	1281	359	0.0965	-	7470A	PPB	29 Feb 2024 18:58:44	U	SMPL
CCV03 - 1	CCV03	22496	4.7732	-	7470A	PPB	29 Feb 2024 19:01:01	U	SMPL
CCB03 - 1	CCB03	-59	0.0082	-	7470A	PPB	29 Feb 2024 19:03:17	U	SMPL
PB159266TB - 1	PB159266TB	10	0.0228	-	7470A	PPB	29 Feb 2024 19:05:35	U	SMPL
P1585-02LX5 - 1		-29	0.0145	-	7470A	PPB	29 Feb 2024 19:07:51	U	SMPL
P1585-02A - 1		18539	3.9373	-	7470A	PPB	29 Feb 2024 19:10:06	U	SMPL
P1587-09LX5 - 1		218	0.0667	-	7470A	PPB	29 Feb 2024 19:12:22	U	SMPL
P1587-09A - 1		21669	4.5985	-	7470A	PPB	29 Feb 2024 19:14:40	U	SMPL
CCV04 - 1	CCV04	24002	5.0914	-	7470A	PPB	29 Feb 2024 19:16:56	U	SMPL
CCB04 - 1	CCB04	-92	0.0012	-	7470A	PPB	29 Feb 2024 19:19:14	U	SMPL