

LB129784


SFAM01.1

INSTRUMENT ID: CV

Linear

 μ Abs.:

50386



A= 0.0000e+000

B= 1.9726e-004 slope

C= 3.3000e-002 y-intercept

Rho= 0.9982932

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.00	0.000	0.079	0.079	234	0.000	234					5
0.05	0.050					442	0				2
0.20	0.200	0.197	-0.003	829	0.0 %	829					6
2.50	2.500	2.647	0.147	13250	0.0 %	13250					9
5.00	5.000	4.563	-0.437	22964	0.0 %	22964					3
7.50	7.500	7.743	0.243	39085	0.0 %	39085					6
10.0	10.000	9.972	-0.028	50386	0.0 %	50386					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	234	-		0 SFAM01.1	PPB	2/29/2024 20:13	Std
	0.2 S01	829	-		0.2 SFAM01.1	PPB	2/29/2024 20:15	Std
	2.5 S02	13250	-		2.5 SFAM01.1	PPB	2/29/2024 20:18	Std
	5 S03	22964	-		5 SFAM01.1	PPB	2/29/2024 20:20	Std
	7.5 S04	39085	-		7.5 SFAM01.1	PPB	2/29/2024 20:22	Std
	10 S05	50386	-		10 SFAM01.1	PPB	2/29/2024 20:27	Std
ICV027	ICV027	20475	4.0718 -		SFAM01.1	PPB	2/29/2024 20:30	SMPL
ICB027	ICB027	24	0.0377 -		SFAM01.1	PPB	2/29/2024 20:33	SMPL
CCV005	CCV005	22979	4.5658 -		SFAM01.1	PPB	2/29/2024 20:35	SMPL
CCB005	CCB005	24	0.0377 -		SFAM01.1	PPB	2/29/2024 20:37	SMPL
PB159378BL	PBW378	65	0.0458 -		SFAM01.1	PPB	2/29/2024 20:39	SMPL
P1466-01	MCOAB5	360	0.104 -		SFAM01.1	PPB	2/29/2024 20:42	SMPL
P1466-02	MCOAE3	317	0.0955 -		SFAM01.1	PPB	2/29/2024 20:44	SMPL
P1466-03	MCOAF0	377	0.1074 -		SFAM01.1	PPB	2/29/2024 20:46	SMPL
P1466-04	MCOAF0D	365	0.105 -		SFAM01.1	PPB	2/29/2024 20:48	SMPL
P1466-05	MCOAF0S	5447	1.1075 -		SFAM01.1	PPB	2/29/2024 20:51	SMPL
P1466-06	MCOAF1	322	0.0965 -		SFAM01.1	PPB	2/29/2024 20:53	SMPL
P1466-07	MCOAE6	299	0.092 -		SFAM01.1	PPB	2/29/2024 20:55	SMPL
P1466-08	MCOAF2	282	0.0886 -		SFAM01.1	PPB	2/29/2024 20:58	SMPL
P1466-09	MCOAG0	245	0.0813 -		SFAM01.1	PPB	2/29/2024 21:00	SMPL
P1466-10	MCOAG3	223	0.077 -		SFAM01.1	PPB	2/29/2024 21:02	SMPL
P1466-11	MCOAD5	1583	0.3453 -		SFAM01.1	PPB	2/29/2024 21:04	SMPL
P1466-12	MCOAD6	526	0.1368 -		SFAM01.1	PPB	2/29/2024 21:07	SMPL
P1466-13	MCOAE2	243	0.0809 -		SFAM01.1	PPB	2/29/2024 21:09	SMPL
P1466-14	MCOAE5	304	0.093 -		SFAM01.1	PPB	2/29/2024 21:11	SMPL
P1466-15	MCOAE8	313	0.0947 -		SFAM01.1	PPB	2/29/2024 21:13	SMPL
P1466-16	MCOAG1	324	0.0969 -		SFAM01.1	PPB	2/29/2024 21:16	SMPL
P1466-17	MCOAG4	352	0.1024 -		SFAM01.1	PPB	2/29/2024 21:18	SMPL
P1466-18	MCOAB2	457	0.1231 -		SFAM01.1	PPB	2/29/2024 21:20	SMPL
P1466-19	MCOAB3	372	0.1064 -		SFAM01.1	PPB	2/29/2024 21:22	SMPL
CCV006	CCV006	22865	4.5433 -		SFAM01.1	PPB	2/29/2024 21:25	SMPL
CCB006	CCB006	58	0.0444 -		SFAM01.1	PPB	2/29/2024 21:27	SMPL
P1466-20	MCOAD2	511	0.1338 -		SFAM01.1	PPB	2/29/2024 21:29	SMPL
P1466-21	MCOAG2	180	0.0685 -		SFAM01.1	PPB	2/29/2024 21:32	SMPL
P1466-22	MCOAG5	147	0.062 -		SFAM01.1	PPB	2/29/2024 21:34	SMPL
PB159379BL	PBW379	17	0.0364 -		SFAM01.1	PPB	2/29/2024 21:36	SMPL
P1499-01	MCOAB6	247	0.0817 -		SFAM01.1	PPB	2/29/2024 21:38	SMPL
P1499-02	MCOAB6D	237	0.0798 -		SFAM01.1	PPB	2/29/2024 21:41	SMPL
P1499-03	MCOAB6S	5645	1.1465 -		SFAM01.1	PPB	2/29/2024 21:43	SMPL
P1499-04	MCOAC4	123	0.0573 -		SFAM01.1	PPB	2/29/2024 21:45	SMPL
P1499-05	MCOAH5	19796	3.9379 -		SFAM01.1	PPB	2/29/2024 21:47	SMPL
P1499-06	MCOAB7	594	0.1502 -		SFAM01.1	PPB	2/29/2024 21:50	SMPL
P1499-07	MCOAC2	195	0.0715 -		SFAM01.1	PPB	2/29/2024 21:52	SMPL
P1499-08	MCOAE1	135	0.0596 -		SFAM01.1	PPB	2/29/2024 21:54	SMPL
P1499-09	MCOAB4	162	0.065 -		SFAM01.1	PPB	2/29/2024 21:57	SMPL
P1499-10	MCOAC6	422	0.1162 -		SFAM01.1	PPB	2/29/2024 21:59	SMPL

LB129734

INSTRUMENT ID : CV1

P1499-11	MC0AC7	266	0.0855 -	SFAM01.1	PPB	2/29/2024 22:01 SMPL
P1499-12	MC0AE7	186	0.0697 -	SFAM01.1	PPB	2/29/2024 22:03 SMPL
CCV007	CCV007	23725	4.7129 -	SFAM01.1	PPB	2/29/2024 22:06 SMPL
CCB007	CCB007	80	0.0488 -	SFAM01.1	PPB	2/29/2024 22:08 SMPL