

LB129738


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49922



A= 0.0000e+000

B= 1.9874e-004 slope

C= -1.0018e-003 y-intercept

Rho= 0.9996090

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/d
0.00	0.000	0.069	0.069	350	0.000	350					=
0.05	0.050					442	0				0
0.20	0.200	0.200	0.000	1012	0.0 %	1012					-3
2.50	2.500	2.420	-0.080	12181	0.0 %	12181					-2
5.00	5.000	4.900	-0.100	24662	0.0 %	24662					3
7.50	7.500	7.691	0.191	38704	0.0 %	38704					-1
10.0	10.000	9.920	-0.080	49922	0.0 %	49922					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	350	-		0 SFAM01.1	PPB	3/1/2024 10:14	Std
	0.2 S01	1012	-		0.2 SFAM01.1	PPB	3/1/2024 10:17	Std
	2.5 S02	12181	-		2.5 SFAM01.1	PPB	3/1/2024 10:20	Std
	5 S03	24662	-		5 SFAM01.1	PPB	3/1/2024 10:23	Std
	7.5 S04	38704	-		7.5 SFAM01.1	PPB	3/1/2024 10:25	Std
	10 S05	49922	-		10 SFAM01.1	PPB	3/1/2024 10:30	Std
ICV028	ICV028	20254	4.0242 -		SFAM01.1	PPB	3/1/2024 10:33	SMPL
ICB028	ICB028	108	0.0205 -		SFAM01.1	PPB	3/1/2024 10:35	SMPL
CCV008	CCV008	24774	4.9225 -		SFAM01.1	PPB	3/1/2024 10:37	SMPL
CCB008	CCB008	53	0.0095 -		SFAM01.1	PPB	3/1/2024 10:40	SMPL
PB159391BL	PBS391	77	0.0143 -		SFAM01.1	PPB	3/1/2024 10:42	SMPL
P1416-01	CODZ6	8995	1.7866 -		SFAM01.1	PPB	3/1/2024 10:44	SMPL
P1416-02	CODZ6D	8929	1.7735 -		SFAM01.1	PPB	3/1/2024 10:46	SMPL
P1416-03	CODZ6S	22618	4.494 -		SFAM01.1	PPB	3/1/2024 10:49	SMPL
P1518-01	MBH9R9	2714	0.5384 -		SFAM01.1	PPB	3/1/2024 10:51	SMPL
P1518-02	MBH9R9D	2442	0.4843 -		SFAM01.1	PPB	3/1/2024 10:53	SMPL
P1518-03	MBH9R9S	16718	3.3215 -		SFAM01.1	PPB	3/1/2024 10:56	SMPL
PB159392BL	PBS392	121	0.023 -		SFAM01.1	PPB	3/1/2024 10:58	SMPL
P1443-01	MBH9Q7	1283	0.254 -		SFAM01.1	PPB	3/1/2024 11:00	SMPL
P1443-02	MBH9Q8	1563	0.3096 -		SFAM01.1	PPB	3/1/2024 11:02	SMPL
P1443-03	MBH9T6	1075	0.2126 -		SFAM01.1	PPB	3/1/2024 11:05	SMPL
P1443-04	MBH9T7	1598	0.3166 -		SFAM01.1	PPB	3/1/2024 11:07	SMPL
P1443-06	MBH9Q9	1821	0.3609 -		SFAM01.1	PPB	3/1/2024 11:09	SMPL
P1443-07	MBH9Q9D	1854	0.3675 -		SFAM01.1	PPB	3/1/2024 11:11	SMPL
P1443-08	MBH9Q9S	15931	3.1651 -		SFAM01.1	PPB	3/1/2024 11:14	SMPL
P1443-09	MBH9T8	490	0.0964 -		SFAM01.1	PPB	3/1/2024 11:16	SMPL
P1443-10	MBH9T9	466	0.0916 -		SFAM01.1	PPB	3/1/2024 11:18	SMPL
P1443-11	MBH9W0	5371	1.0664 -		SFAM01.1	PPB	3/1/2024 11:21	SMPL
P1443-12	MBH9W1	1346	0.2665 -		SFAM01.1	PPB	3/1/2024 11:23	SMPL
P1443-13	MBH9Q6	3316	0.658 -		SFAM01.1	PPB	3/1/2024 11:25	SMPL
CCV009	CCV009	24632	4.8943 -		SFAM01.1	PPB	3/1/2024 11:27	SMPL
CCB009	CCB009	120	0.0228 -		SFAM01.1	PPB	3/1/2024 11:30	SMPL
P1443-14	MBH9S0	2793	0.5541 -		SFAM01.1	PPB	3/1/2024 11:32	SMPL
P1443-15	MBH9T4	2798	0.5551 -		SFAM01.1	PPB	3/1/2024 11:34	SMPL
P1443-16	MBH9T5	3412	0.6771 -		SFAM01.1	PPB	3/1/2024 11:36	SMPL
P1443-17	MBH9W2	1103	0.2182 -		SFAM01.1	PPB	3/1/2024 11:39	SMPL
P1443-18	MBH9W3	522	0.1027 -		SFAM01.1	PPB	3/1/2024 11:41	SMPL
P1443-20	MBH9R1	745	0.1471 -		SFAM01.1	PPB	3/1/2024 11:43	SMPL
P1443-21	MBH9R2	1393	0.2758 -		SFAM01.1	PPB	3/1/2024 11:46	SMPL
P1443-22	MBH9R3	2401	0.4762 -		SFAM01.1	PPB	3/1/2024 11:48	SMPL
CCV010	CCV010	24583	4.8845 -		SFAM01.1	PPB	3/1/2024 11:50	SMPL
CCB010	CCB010	117	0.0223 -		SFAM01.1	PPB	3/1/2024 11:52	SMPL