



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 3/4/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 03/01/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 03/02/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB129746

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1660-08	BP-VPB-195-GW-1000-1002	1	1.15	8.83	9.98	1.89	8.4	sludge sample
P1661-05	SVOC-GPC-BLANK	2	1.00	1.00	2.00	2.00	100.0	
P1661-06	PEST-GPC-BLANK	3	1.00	1.00	2.00	2.00	100.0	
P1661-07	PEST-GPC-BLANK-SPIKE	4	1.00	1.00	2.00	2.00	100.0	
P1661-08	PCB-GPC-BLANK	5	1.00	1.00	2.00	2.00	100.0	
P1661-09	PCB-GPC-BLANK-SPIKE	6	1.00	1.00	2.00	2.00	100.0	
P1661-10	SVOC-GPC2-BLANK	7	1.00	1.00	2.00	2.00	100.0	
P1661-11	PEST-GPC2-BLANK	8	1.00	1.00	2.00	2.00	100.0	
P1661-12	PEST-GPC2-BLANK-SPIKE	9	1.00	1.00	2.00	2.00	100.0	
P1661-13	PCB-GPC2-BLANK-SPIKE	10	1.00	1.00	2.00	2.00	100.0	
P1661-14	PCB-GCP2-BLANK-SPIKE	11	1.00	1.00	2.00	2.00	100.0	
P1662-01	ETGI-352	12	1.12	8.76	9.88	8.67	86.2	
P1662-03	RBR-2520556	13	1.15	8.41	9.56	8.37	85.9	
P1666-01	OR-02-030124	20	1.18	8.56	9.74	9.03	91.7	
P1666-02	OR-02-030124-E2	21	1.15	8.84	9.99	9.41	93.4	
P1669-01	TP-14	14	1.15	8.83	9.98	8.8	86.6	
P1669-02	TP-14-EPH	15	1.19	8.67	9.86	8.54	84.8	
P1669-05	TP-11	16	1.15	8.83	9.98	9.13	90.4	
P1669-06	TP-11-EPH	17	1.16	8.47	9.63	8.47	86.3	
P1669-09	TP-6	18	1.14	8.84	9.98	8.54	83.7	
P1669-10	TP-6-EPH	19	1.17	8.53	9.7	8.08	81.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$