



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129750

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	03/01/2024	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	03/01/2024	15:01
3	CAL3	1	Water	NA	NA	20.2	20.3	03/01/2024	15:05
4	ICV	1	Water	NA	NA	20.5	7.00	03/01/2024	15:10
5	CCV1	1	Water	NA	NA	20.3	2.01	03/01/2024	15:15
6	P1659-04	1	Solid	20.02	20	23.1	5.95	03/01/2024	15:20
7	P1659-08	1	Solid	20.03	20	22.5	5.48	03/01/2024	15:25
8	P1662-02	1	Solid	20.05	20	24.1	8.32	03/01/2024	15:40
9	P1662-04	1	Solid	20.02	20	24.1	8.45	03/01/2024	15:50
10	P1669-04	1	Solid	20.01	20	24.5	7.28	03/01/2024	16:00
11	P1669-08	1	Solid	20.02	20	24.0	7.02	03/01/2024	16:10
12	P1669-12	1	Solid	20.03	20	24.3	7.09	03/01/2024	16:15
13	P1669-12DUP	1	Solid	20.04	20	24.4	7.10	03/01/2024	16:16
14	CCV2	1	Water	NA	NA	20.4	12.02	03/01/2024	16:20

WORKLIST(Hardcopy Internal Chain)

VB 1201450

WorkList Name : corrosivity p1662 WorkList ID : 178335 Department : Wet-Chemistry Date : 03-01-2024 14:21:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1659-04	TP-16	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/29/2024	9045D
P1659-08	TP-7	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/29/2024	9045D
P1662-02	ETGI-352	Solid	Corrosivity	Cool 4 deg C	PSEG03	H13	03/01/2024	9045D
P1662-04	RBR-2520556	Solid	Corrosivity	Cool 4 deg C	PSEG03	H13	03/01/2024	9045D
P1669-04	TP-14	Solid	Corrosivity	Cool 4 deg C	PSEG03	H13	03/01/2024	9045D
P1669-08	TP-11	Solid	Corrosivity	Cool 4 deg C	PSEG03	H13	03/01/2024	9045D
P1669-12	TP-6	Solid	Corrosivity	Cool 4 deg C	PSEG03	H13	03/01/2024	9045D

Date/Time 03/01/24 14:30
 Raw Sample Received by: RD WCC
 Raw Sample Relinquished by: JDCSM

Date/Time 03/01/24 17:00
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: RD WCC