

LB129751


SFAM01.1

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

53643



A= 0.0000e+000

B= 1.8881e-004 slope

C= -8.7731e-003 y-intercept

Rho= 0.9994995

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.00	0.000	0.006	0.006	77	0.000	77					=
0.05	0.050					442	0				-1
0.20	0.200	0.198	-0.002	1097	0.0 %	1097					1
2.50	2.500	2.523	0.023	13408	0.0 %	13408					2
5.00	5.000	5.094	0.094	27027	0.0 %	27027					-3
7.50	7.500	7.259	-0.241	38493	0.0 %	38493					1
10.0	10.000	10.120	0.120	53643	0.0 %	53643					

LB129751

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	77	-		0 SFAM01.1	PPB	3/1/2024 16:04	Std
	0.2 S01	1097	-		0.2 SFAM01.1	PPB	3/1/2024 16:06	Std
	2.5 S02	13408	-		2.5 SFAM01.1	PPB	3/1/2024 16:09	Std
	5 S03	27027	-		5 SFAM01.1	PPB	3/1/2024 16:11	Std
	7.5 S04	38493	-		7.5 SFAM01.1	PPB	3/1/2024 16:13	Std
	10 S05	53643	-		10 SFAM01.1	PPB	3/1/2024 16:15	Std
ICV030	ICV030	21147	3.984 -		SFAM01.1	PPB	3/1/2024 16:18	SMPL
ICB030	ICB030	-119	-0.0312 -		SFAM01.1	PPB	3/1/2024 16:20	SMPL
CCV015	CCV015	27675	5.2166 -		SFAM01.1	PPB	3/1/2024 16:23	SMPL
CCB015	CCB015	-155	-0.038 -		SFAM01.1	PPB	3/1/2024 16:25	SMPL
PB159405BL	PBW405	24	-0.0042 -		SFAM01.1	PPB	3/1/2024 16:27	SMPL
P1417-01	CODY6	44	-0.0005 -		SFAM01.1	PPB	3/1/2024 16:30	SMPL
P1417-02	CODY7	55	0.0016 -		SFAM01.1	PPB	3/1/2024 16:32	SMPL
P1417-03	CODY8	51	0.0009 -		SFAM01.1	PPB	3/1/2024 16:34	SMPL
P1417-04	CODY9	47	0.0001 -		SFAM01.1	PPB	3/1/2024 16:36	SMPL
P1417-05	CODY9D	46	-0.0001 -		SFAM01.1	PPB	3/1/2024 16:39	SMPL
P1417-06	CODY9S	5710	1.0693 -		SFAM01.1	PPB	3/1/2024 16:41	SMPL
P1417-08	CODQ5	9310	1.7491 -		SFAM01.1	PPB	3/1/2024 16:45	SMPL
P1443-05	MBHA07	-74	-0.0227 -		SFAM01.1	PPB	3/1/2024 16:50	SMPL
P1443-19	MBHA08	26	-0.0039 -		SFAM01.1	PPB	3/1/2024 16:52	SMPL
CCV016	CCV016	26463	4.9878 -		SFAM01.1	PPB	3/1/2024 16:55	SMPL
CCB016	CCB016	-197	-0.046 -		SFAM01.1	PPB	3/1/2024 16:57	SMPL