



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129761

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/04/2024	12:30
2	CAL2	1	Water	NA	NA	20.3	7.00	03/04/2024	12:31
3	CAL3	1	Water	NA	NA	20.3	10.02	03/04/2024	12:33
4	ICV	1	Water	NA	NA	20.3	7.01	03/04/2024	12:35
5	CCV1	1	Water	NA	NA	20.3	2.01	03/04/2024	12:40
6	P1670-01	1	Water	NA	NA	24.3	4.39	03/04/2024	12:45
7	P1670-01DUP	1	Water	NA	NA	24.4	4.40	03/04/2024	12:46
8	P1671-01	1	Water	NA	NA	23.4	4.23	03/04/2024	12:50
9	P1671-02	1	Water	NA	NA	23.6	4.36	03/04/2024	12:55
10	CCV2	1	Water	NA	NA	20.3	12.02	03/04/2024	15:00

VB 129861

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1670 water WorkList ID : 178360 Department : Wet-Chemistry Date : 03-04-2024 12:14:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1670-01	17 GAMMA-JARINU	Water	pH	Cool 4 deg C	STER18	H21	02/26/2024	9040C
P1671-01	2 JS9600	Water	pH	Cool 4 deg C	STER18	H21	02/27/2024	9040C
P1671-02	15 JS7500	Water	pH	Cool 4 deg C	STER18	H21	02/27/2024	9040C

Date/Time 03/04/24 12:20
Raw Sample Received by: Jol Cox
Raw Sample Relinquished by: Jol Cox

Date/Time 03/04/24 15:00
Raw Sample Received by: Jol Cox
Raw Sample Relinquished by: Jol Cox