



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129762

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/04/2024	12:30
2	CAL2	1	Water	NA	NA	20.3	7.00	03/04/2024	12:31
3	CAL3	1	Water	NA	NA	20.3	10.02	03/04/2024	12:33
4	ICV	1	Water	NA	NA	20.3	7.01	03/04/2024	12:35
5	CCV1	1	Water	NA	NA	20.4	2.02	03/04/2024	14:00
6	P1673-01	1	Solid	20.02	20	24.6	6.66	03/04/2024	14:10
7	P1673-01DUP	1	Solid	20.03	20	24.8	6.67	03/04/2024	14:11
8	P1674-05	1	Solid	20.01	20	24.7	12.72	03/04/2024	14:30
9	CCV2	1	Water	NA	NA	20.3	12.02	03/04/2024	14:33

VB 129462

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1673 solids WorkList ID : 178364 Department : Wet-Chemistry Date : 03-04-2024 13:17:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1673-01	RT-2321	Solid	pH	Cool 4 deg C	PSEG03	H21	03/04/2024	9045D
P1674-05	PAINT-CHIPS	Solid	pH	Cool 4 deg C	PSEG03	H21	03/04/2024	9045D

Date/Time 03/04/24 13:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 03/04/24 16:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]