



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129764

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/04/2024	12:40
2	CAL2	1	Water	NA	NA	20.3	6.99	03/04/2024	12:41
3	CAL3	1	Water	NA	NA	20.3	10.02	03/04/2024	12:45
4	ICV	1	Water	NA	NA	20.4	7.02	03/04/2024	12:50
5	CCV1	1	Water	NA	NA	20.4	2.02	03/04/2024	12:55
6	P1672-04	1	Solid	20.02	20	25.5	7.82	03/04/2024	13:00
7	P1672-04DUP	1	Solid	20.03	20	25.6	7.83	03/04/2024	13:01
8	P1672-08	1	Solid	20.01	20	24.3	6.02	03/04/2024	13:10
9	P1672-12	1	Solid	20.03	20	23.3	7.61	03/04/2024	13:15
10	P1672-16	1	Solid	20.04	20	23.4	7.31	03/04/2024	13:20
11	CCV2	1	Water	NA	NA	20.3	12.02	03/04/2024	13:25

129764

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1672 WorkList ID : 178363 Department : Wet-Chemistry Date : 03-04-2024 12:15:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1672-04	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	03/02/2024	9045D
P1672-08	TP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	03/01/2024	9045D
P1672-12	TP-5	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	03/02/2024	9045D
P1672-16	TP-13	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	03/02/2024	9045D

Date/Time 03/04/24 12:30
Raw Sample Received by: JOL
Raw Sample Relinquished by: JOL

Date/Time 03/04/24 16:00
Raw Sample Received by: JOL
Raw Sample Relinquished by: JOL