

LB129775


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

59233



A= 0.0000e+000

B= 1.6806e-004 slope

C= 2.7721e-002 intercept

Rho= 0.9999106

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.042	0.042	82	0.000	82					0
0.05	0.050					442	0				1
0.20	0.200	0.220	0.020	1146	0.0 %	1146					10
2.50	2.500	2.414	-0.086	14200	0.0 %	14200					3
5.00	5.000	4.976	-0.024	29444	0.0 %	29444					0
7.50	7.500	7.565	0.065	44849	0.0 %	44849					0
10.0	10.000	9.983	-0.017	59233	0.0 %	59233					0

LB129775

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	82	-		0 SFAM01.1	PPB	3/5/2024 13:42	Std
	0.2 S01	1146	-		0.2 SFAM01.1	PPB	3/5/2024 13:44	Std
	2.5 S02	14200	-		2.5 SFAM01.1	PPB	3/5/2024 13:46	Std
	5 S03	29444	-		5 SFAM01.1	PPB	3/5/2024 13:49	Std
	7.5 S04	44849	-		7.5 SFAM01.1	PPB	3/5/2024 13:51	Std
	10 S05	59233	-		10 SFAM01.1	PPB	3/5/2024 13:56	Std
ICV033	ICV033	23347	3.9515 -		SFAM01.1	PPB	3/5/2024 13:59	SMPL
ICB033	ICB033	-104	0.0102 -		SFAM01.1	PPB	3/5/2024 14:01	SMPL
CCV025	CCV025	29816	5.0387 -		SFAM01.1	PPB	3/5/2024 14:03	SMPL
CCB025	CCB025	-93	0.0121 -		SFAM01.1	PPB	3/5/2024 14:05	SMPL
PB159457BL	PBW457	142	0.0516 -		SFAM01.1	PPB	3/5/2024 14:08	SMPL
P1497-01	MBHA09	156	0.0539 -		SFAM01.1	PPB	3/5/2024 14:10	SMPL
P1497-02	MBHA10	162	0.0549 -		SFAM01.1	PPB	3/5/2024 14:12	SMPL
P1497-03	MBHA11	186	0.059 -		SFAM01.1	PPB	3/5/2024 14:15	SMPL
P1497-04	MBHA26	265	0.0723 -		SFAM01.1	PPB	3/5/2024 14:17	SMPL
P1497-05	MBHA26D	313	0.0803 -		SFAM01.1	PPB	3/5/2024 14:22	SMPL
P1497-06	MBHA26S	5277	0.9146 -		SFAM01.1	PPB	3/5/2024 14:25	SMPL
P1497-07	MBHA27	485	0.1092 -		SFAM01.1	PPB	3/5/2024 14:27	SMPL
P1497-08	MBHA03	629	0.1334 -		SFAM01.1	PPB	3/5/2024 14:29	SMPL
P1497-09	MBHA05	512	0.1138 -		SFAM01.1	PPB	3/5/2024 14:32	SMPL
P1497-10	MBH9Y0	2708	0.4828 -		SFAM01.1	PPB	3/5/2024 14:34	SMPL
P1497-11	MBHA06	1216	0.2321 -		SFAM01.1	PPB	3/5/2024 14:36	SMPL
P1497-12	MBHA25	615	0.1311 -		SFAM01.1	PPB	3/5/2024 14:38	SMPL
P1497-13	MBHA28	35	0.0336 -		SFAM01.1	PPB	3/5/2024 14:41	SMPL
P1550-01	MCOAF4	296	0.0775 -		SFAM01.1	PPB	3/5/2024 14:45	SMPL
P1550-02	MCOAF5	336	0.0842 -		SFAM01.1	PPB	3/5/2024 14:47	SMPL
P1550-03	MCOAF6	202	0.0617 -		SFAM01.1	PPB	3/5/2024 14:50	SMPL
P1550-04	MCOAF7	69	0.0393 -		SFAM01.1	PPB	3/5/2024 14:52	SMPL
P1550-05	MCOAF9	108	0.0459 -		SFAM01.1	PPB	3/5/2024 14:54	SMPL
CCV026	CCV026	28284	4.7812 -		SFAM01.1	PPB	3/5/2024 14:56	SMPL
CCB026	CCB026	-121	0.0074 -		SFAM01.1	PPB	3/5/2024 14:59	SMPL
P1550-06	MCOAB9	1595	0.2958 -		SFAM01.1	PPB	3/5/2024 15:01	SMPL
P1550-07	MCOACO	27	0.0323 -		SFAM01.1	PPB	3/5/2024 15:03	SMPL
P1550-08	MCOADO	1126	0.217 -		SFAM01.1	PPB	3/5/2024 15:06	SMPL
P1550-09	MCOAD0D	1123	0.2165 -		SFAM01.1	PPB	3/5/2024 15:08	SMPL
P1550-10	MCOAD0S	6127	1.0574 -		SFAM01.1	PPB	3/5/2024 15:10	SMPL
P1550-11	MCOAD8	-52	0.019 -		SFAM01.1	PPB	3/5/2024 15:12	SMPL
P1550-12	MCOAF8	50	0.0361 -		SFAM01.1	PPB	3/5/2024 15:15	SMPL
PB159458BL	PBW458	25	0.0319 -		SFAM01.1	PPB	3/5/2024 15:17	SMPL
CCV027	CCV027	28177	4.7632 -		SFAM01.1	PPB	3/5/2024 15:19	SMPL
CCB027	CCB027	-127	0.0064 -		SFAM01.1	PPB	3/5/2024 15:22	SMPL