

LB129782


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43093



A= 0.0000e+000

B= 2.3350e-004

C= 1.2449e-002

Rho= 0.9999156

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	0.028	0.028	66	0.000	66	0				12
0.2	0.200	0.224	0.024	907	0.0 %	907					11
2.5	2.500	2.480	-0.020	10566	0.0 %	10566					11
5.0	5.000	4.967	-0.033	21217	0.0 %	21217					11
7.5	7.500	7.427	-0.073	31754	0.0 %	31754					11
10.0	10.000	10.075	0.075	43093	0.0 %	43093					1

LB129782 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Mar 2024 16:03:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	66	-	0.0000	7470A	PPB	05 Mar 2024 16:04:35	S	Std
0.2 - 1	S0.2	907	-	0.2000	7470A	PPB	05 Mar 2024 16:06:54	S	Std
2.5 - 1	S2.5	10566	-	2.5000	7470A	PPB	05 Mar 2024 16:09:12	S	Std
5.0 - 1	S5	21217	-	5.0000	7470A	PPB	05 Mar 2024 16:11:29	S	Std
7.5 - 1	S7.5	31754	-	7.5000	7470A	PPB	05 Mar 2024 16:13:47	S	Std
10.0 - 1	S10	43093	-	10.0000	7470A	PPB	05 Mar 2024 16:21:17	S	Std
ICV34 - 1	ICV34	15483	3.6277	-	7470A	PPB	05 Mar 2024 16:24:06	U	SMPL
ICB34 - 1	ICB34	-125	-0.0167	-	7470A	PPB	05 Mar 2024 16:26:21	U	SMPL
CCV28 - 1	CCV28	21321	4.9909	-	7470A	PPB	05 Mar 2024 16:28:41	U	SMPL
CCB28 - 1	CCB28	-137	-0.0195	-	7470A	PPB	05 Mar 2024 16:30:56	U	SMPL
CRA - 1	CRA	849	0.2107	-	7470A	PPB	05 Mar 2024 16:33:13	U	SMPL
High Std - 1	High Std	43468	10.1622	-	7470A	PPB	05 Mar 2024 16:35:28	U	SMPL
CHKstd - 1	CHKstd	29471	6.8939	-	7470A	PPB	05 Mar 2024 16:37:44	U	SMPL
PBW - 1	PBW	-225	-0.0401	-	7470A	PPB	05 Mar 2024 16:40:03	U	SMPL
LCSW - 1	LCSW	15648	3.6663	-	7470A	PPB	05 Mar 2024 16:42:21	U	SMPL
O4699-09 - 1	MDL-WATER-03-QT4-2023	164	0.0507	-	7470A	PPB	05 Mar 2024 16:49:45	U	SMPL
CCV29 - 1	CCV29	21360	5.0000	-	7470A	PPB	05 Mar 2024 16:52:01	U	SMPL
CCB29 - 1	CCB29	-111	-0.0135	-	7470A	PPB	05 Mar 2024 16:54:18	U	SMPL