



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129794

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	03/06/2024	10:00
2	P1665-01		1	12	100.00	>212.0	>212.0	03/06/2024	10:30
3	P1665-01DUP		1	13	100.00	>212.0	>212.0	03/06/2024	11:00
4	P1683-03		1	14	100.00	>212.0	>212.0	03/06/2024	11:30
5	P1683-04		1	13	100.00	>212.0	>212.0	03/06/2024	12:30
6	P1683-05		1	12	100.00	>212.0	>212.0	03/06/2024	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06