



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129797

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/06/2024	16:25
2	CAL2	1	Water	NA	NA	20.3	7.00	03/06/2024	16:26
3	CAL3	1	Water	NA	NA	20.3	10.02	03/06/2024	16:29
4	ICV	1	Water	NA	NA	20.4	7.00	03/06/2024	16:33
5	CCV1	1	Water	NA	NA	20.3	2.01	03/06/2024	16:37
6	P1683-03	1	Water	NA	NA	23.2	12.83	03/06/2024	16:45
7	P1692-01	1	Water	NA	NA	20.3	6.47	03/06/2024	16:52
8	P1693-02	1	Water	NA	NA	20.4	8.26	03/06/2024	17:00
9	P1693-02DUP	1	Water	NA	NA	20.5	8.27	03/06/2024	17:01
10	CCV2	1	Water	NA	NA	20.3	12.02	03/06/2024	17:05

WORKLIST(Hardcopy Internal Chain)

LB120797

WorkList Name : PH P1693

WorkList ID : 178428

Department : Wet-Chemistry

Date : 03-06-2024 16:18:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1683-03	C 3524	Water	pH	Cool 4 deg C	PSEG03	E11	03/05/2024	9040C
P1692-01	A 34745-34746	Water	pH	Cool 4 deg C	PSEG03	I11	03/06/2024	9040C
P1693-02	A 60182	Water	pH	Cool 4 deg C	PSEG03	I11	03/06/2024	9040C

Date/Time 03/06/24 16:20
 Raw Sample Received by: JH 1000
 Raw Sample Relinquished by: JDCSM

Date/Time 03/06/24 17:40
 Raw Sample Received by: JH 1000
 Raw Sample Relinquished by: JH 1000