

LB129801


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

48426



A= 0.0000e+000

B= 2.0342e-004 slope

C= -2.2581e-002 *intercept*

Rho= 0.9993868

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.015	-0.015	39	0.000	39				
0.05	0.050					442	0			
0.20	0.200	0.183	-0.017	1012	0.0 %	1012				
2.50	2.500	2.514	0.014	12469	0.0 %	12469				
5.00	5.000	4.933	-0.067	24363	0.0 %	24363				
7.50	7.500	7.756	0.256	38239	0.0 %	38239				
10.0	10.000	9.828	-0.172	48426	0.0 %	48426				

Handwritten notes:
 9/11
 =
 -149
 1
 -1
 3
 >+ -2
 0.8
 3/22/24

LB129801

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	39	-		0 SFAM01.1	PPB	3/7/2024 10:07	Std
	0.2 S01	1012	-		0.2 SFAM01.1	PPB	3/7/2024 10:10	Std
	2.5 S02	12469	-		2.5 SFAM01.1	PPB	3/7/2024 10:12	Std
	5 S03	24363	-		5 SFAM01.1	PPB	3/7/2024 10:14	Std
	7.5 S04	38239	-		7.5 SFAM01.1	PPB	3/7/2024 10:16	Std
	10 S05	48426	-		10 SFAM01.1	PPB	3/7/2024 10:20	Std
ICV038	ICV038	.19530	3.9502 -		SFAM01.1	PPB	3/7/2024 10:24	SMPL
ICB038	ICB038	-174	-0.058 -		SFAM01.1	PPB	3/7/2024 10:26	SMPL
CCV039	CCV039	25092	5.0816 -		SFAM01.1	PPB	3/7/2024 10:29	SMPL
CCB039	CCB039	-224	-0.0681 -		SFAM01.1	PPB	3/7/2024 10:31	SMPL
PB159496BL	PBW496	-15	-0.0256 -		SFAM01.1	PPB	3/7/2024 10:33	SMPL
P1680-01	E0705	304	0.0393 -		SFAM01.1	PPB	3/7/2024 10:36	SMPL
CCV040	CCV040	24796	5.0214 -		SFAM01.1	PPB	3/7/2024 10:38	SMPL
CCB040	CCB040	-198	-0.0629 -		SFAM01.1	PPB	3/7/2024 10:40	SMPL