



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129808

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/07/2024	09:00
2	P1692-01		1	12	100.00	>212.0	>212.0	03/07/2024	09:30
3	P1692-01DUP		1	13	100.00	>212.0	>212.0	03/07/2024	10:00
4	P1693-02		1	14	100.00	>212.0	>212.0	03/07/2024	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129808

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/07/24 flashpoint

WorkList ID : 178449

Department : Wet-Chemistry

Date : 03-07-2024 08:23:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1692-01	34745-34746	Water	Flash Point	Cool 4 deg C	PSEG03	I11	03/06/2024	1010B
P1693-02	60182	Water	Flash Point	Cool 4 deg C	PSEG03	I11	03/06/2024	1010B

Date/Time03/07/202408:30

Raw Sample Received by:N.P.W.C

Raw Sample Relinquished by:JN Lee

Date/Time03/07/202411:15

Raw Sample Received by:JN Lee

Raw Sample Relinquished by:N.P.W.C