



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129816

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/08/2024	09:15
2	CAL2	1	Water	NA	NA	20.3	7.00	03/08/2024	09:16
3	CAL3	1	Water	NA	NA	20.3	10.03	03/08/2024	09:20
4	ICV	1	Water	NA	NA	20.3	7.00	03/08/2024	09:22
5	CCV1	1	Water	NA	NA	20.3	2.01	03/08/2024	09:25
6	P1706-02	1	Water	NA	NA	22.3	4.87	03/08/2024	09:35
7	P1706-02DUP	1	Water	NA	NA	22.4	4.88	03/08/2024	09:36
8	P1708-01	1	Water	NA	NA	23.1	6.84	03/08/2024	09:40
9	CCV2	1	Water	NA	NA	20.3	12.02	03/08/2024	09:44

13129816

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1706 WorkList ID : 178466 Department : Wet-Chemistry Date : 03-08-2024 08:58:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1706-02	AUD-1323	Water	pH	Cool 4 deg C	PSEG03	I21	03/07/2024	9040C
P1708-01	TRE-1788	Water	pH	Cool 4 deg C	PSEG03	I22	03/07/2024	9040C

Date/Time 03/08/24 09:10
Raw Sample Received by: JBL
Raw Sample Relinquished by: JBL

Date/Time 03/08/24 11:00
Raw Sample Received by: JBL
Raw Sample Relinquished by: JBL