



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129828

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/08/2024	14:15
2	P1693-01		1	12	100.00	>212.0	>212.0	03/08/2024	14:45
3	P1693-01DUP		1	13	100.00	>212.0	>212.0	03/08/2024	15:15
4	P1698-01		1	8	55.00	131.0	130.0	03/08/2024	15:45
5	P1698-02		1	8	61.00	141.8	140.8	03/08/2024	16:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129828

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint 03/08/24 o WorkList ID : 178471 Department : Wet-Chemistry Date : 03-08-2024 09:30:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1693-01	60187	Solid	Flash Point	Cool 4 deg C	PSEG03	I11	03/06/2024	1010B
P1698-01	NWB-2022	Solid	Flash Point	Cool 4 deg C	PSEG03	I21	03/07/2024	1010B
P1698-02	NWB-2023	Solid	Flash Point	Cool 4 deg C	PSEG03	I21	03/07/2024	1010B

Date/Time 03/08/2024 14:00
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.

Date/Time 03/08/2024 16:50
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.