



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129835

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/11/2024	08:00
2	CAL2	1	Water	NA	NA	20.3	7.00	03/11/2024	08:01
3	CAL3	1	Water	NA	NA	20.3	10.02	03/11/2024	08:05
4	ICV	1	Water	NA	NA	20.5	7.00	03/11/2024	08:10
5	CCV1	1	Water	NA	NA	20.3	2.02	03/11/2024	08:15
6	P1724-01	1	Solid	20.02	20	25.3	7.49	03/11/2024	08:25
7	P1724-02	1	Solid	20.03	20	24.2	6.90	03/11/2024	08:33
8	P1724-03	1	Solid	20.02	20	24.2	8.32	03/11/2024	08:37
9	P1724-03DUP	1	Solid	20.03	20	24.3	8.33	03/11/2024	08:38
10	CCV2	1	Water	NA	NA	20.3	12.02	03/11/2024	08:40

WORKLIST(Hardcopy Internal Chain)

WP 129835

WorkList Name : ph p1734 WorkList ID : 178490 Department : Wet-Chemistry Date : 03-11-2024 07:48:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1724-01	B 13925	Solid	pH	Cool 4 deg C	PSEG03	E21	03/08/2024	9045D
P1724-02	B 10252	Solid	pH	Cool 4 deg C	PSEG03	E21	03/08/2024	9045D
P1724-03	B 90	Solid	pH	Cool 4 deg C	PSEG03	E21	03/08/2024	9045D

Date/Time 03-11-24 07:50
 Raw Sample Received by: JP 1021
 Raw Sample Relinquished by: JP 5m

Date/Time 03-11-24 11:40
 Raw Sample Received by: JP 5m
 Raw Sample Relinquished by: JP 1021