



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129836

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/11/2024	08:00
2	CAL2	1	Water	NA	NA	20.3	7.00	03/11/2024	08:01
3	CAL3	1	Water	NA	NA	20.3	10.02	03/11/2024	08:05
4	ICV	1	Water	NA	NA	20.5	7.00	03/11/2024	08:10
5	CCV1	1	Water	NA	NA	20.3	2.02	03/11/2024	08:15
6	P1724-06	1	Water	NA	NA	20.8	9.62	03/11/2024	09:00
7	P1724-06DUP	1	Water	NA	NA	20.7	9.63	03/11/2024	09:01
8	CCV2	1	Water	NA	NA	20.3	12.02	03/11/2024	09:05

VB 129836

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1724 water WorkList ID : 178491 Department : Wet-Chemistry Date : 03-11-2024 07:49:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1724-06	A 1082	Water	pH	Cool 4 deg C	PSEG03	E21	03/08/2024	9040C

Date/Time 03-11-24 07:55
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 03-11-24 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]