

7470A


LAB 29841

INSTRUMENT ID: CV2

Linear

 μ Abs.:

44776



A= 0.0000e+000

B= 2.2295e-004

C= -1.4104e-002

Rho= 0.9999424

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	b/o 1)
0.0	0.000	-0.013	-0.013	6	0.000	6	0				
0.2	0.200	0.169	-0.031	822	0.0 %	822					-16
2.5	2.500	2.509	0.009	11316	0.0 %	11316					0
5.0	5.000	5.083	0.083	22863	0.0 %	22863					2
7.5	7.500	7.483	-0.017	33624	0.0 %	33624					0
10.0	10.000	9.969	-0.031	44776	0.0 %	44776					6

LB129841 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Mar 2024 14:18:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	6	-	0.0000	7470A	PPB	11 Mar 2024 14:19:06	S	Std
0.2 - 1	50.2	822	-	0.2000	7470A	PPB	11 Mar 2024 14:21:23	S	Std
2.5 - 1	52.5	11316	-	2.5000	7470A	PPB	11 Mar 2024 14:23:40	S	Std
5.0 - 1	55	22863	-	5.0000	7470A	PPB	11 Mar 2024 14:25:57	S	Std
7.5 - 1	57.5	33624	-	7.5000	7470A	PPB	11 Mar 2024 14:28:15	S	Std
10.0 - 1	510	44776	-	10.0000	7470A	PPB	11 Mar 2024 14:30:33	S	Std
ICV43 - 1	ICV43	18007	4.0006	-	7470A	PPB	11 Mar 2024 14:34:17	U	SMPL
ICB43 - 1	ICB43	-187	-0.0558	-	7470A	PPB	11 Mar 2024 14:36:32	U	SMPL
CCV51 - 1	CCV51	22670	5.0403	-	7470A	PPB	11 Mar 2024 14:38:50	U	SMPL
CCB51 - 1	CCB51	-225	-0.0643	-	7470A	PPB	11 Mar 2024 14:41:06	U	SMPL
CRA - 1	CRA	927	0.1926	-	7470A	PPB	11 Mar 2024 14:43:23	U	SMPL
HighStd - 1	HighStd	44480	9.9029	-	7470A	PPB	11 Mar 2024 14:45:38	U	SMPL
ChkStd - 1	ChkStd	31586	7.0281	-	7470A	PPB	11 Mar 2024 14:47:53	U	SMPL
PB159545BL - 1	PBW	-287	-0.0781	-	7470A	PPB	11 Mar 2024 14:50:12	U	SMPL
PB159545BS - 1	LCSW	17781	3.9503	-	7470A	PPB	11 Mar 2024 14:55:10	U	SMPL
P1683-03 - 1	3524	467	0.0900	-	7470A	PPB	11 Mar 2024 14:57:26	U	SMPL
P1686-02 - 1	COMP	1902	0.4100	-	7470A	PPB	11 Mar 2024 14:59:45	U	SMPL
P1706-02 - 1	AUD-1323	3132	0.6842	-	7470A	PPB	11 Mar 2024 15:02:02	U	SMPL
P1708-01 - 1	TRE-1788	-156	-0.0489	-	7470A	PPB	11 Mar 2024 15:04:21	U	SMPL
P1708-01DUP - 1	TRE-1788DUP	7	-0.0125	-	7470A	PPB	11 Mar 2024 15:06:37	U	SMPL
CCV52 - 1	CCV52	21445	4.7672	-	7470A	PPB	11 Mar 2024 15:11:26	U	SMPL
CCB52 - 1	CCB52	-238	-0.0672	-	7470A	PPB	11 Mar 2024 15:13:41	U	SMPL
P1708-01MS - 1	TRE-1788MS	15487	3.4388	-	7470A	PPB	11 Mar 2024 15:17:55	U	SMPL
P1708-01MSD - 1	TRE-1788MSD	15167	3.3675	-	7470A	PPB	11 Mar 2024 15:23:22	U	SMPL
P1708-01LX5 - 1		-171	-0.0522	-	7470A	PPB	11 Mar 2024 15:25:37	U	SMPL
P1708-01A - 1		15258	3.3877	-	7470A	PPB	11 Mar 2024 15:27:55	U	SMPL
CCV53 - 1	CCV53	20563	4.5705	-	7470A	PPB	11 Mar 2024 15:30:11	U	SMPL
CCB53 - 1	CCB53	-305	-0.0821	-	7470A	PPB	11 Mar 2024 15:32:30	U	SMPL