



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129846

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	03/11/2024	11:15
2	P1706-02		1	12	100.00	>212.0	>212.0	03/11/2024	11:45
3	P1706-02DUP		1	11	100.00	>212.0	>212.0	03/11/2024	12:15
4	P1708-01		1	14	100.00	>212.0	>212.0	03/11/2024	12:45
5	P1724-04		1	13	100.00	>212.0	>212.0	03/11/2024	13:15
6	P1724-05		1	14	100.00	>212.0	>212.0	03/11/2024	13:45
7	P1724-07		1	12	100.00	>212.0	>212.0	03/11/2024	14:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129846

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/11/24 flashpoint WorkList ID : 178497 Department : Wet-Chemistry Date : 03-11-2024 08:19:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1706-02	AUD-1323	Water	Flash Point	Cool 4 deg C	PSEG03	I21	03/07/2024	1010B
P1708-01	TRE-1788	Water	Flash Point	Cool 4 deg C	PSEG03	I22	03/07/2024	1010B
P1724-04	380	Water	Flash Point	Cool 4 deg C	PSEG03	E21	03/08/2024	1010B
P1724-05	381	Water	Flash Point	Cool 4 deg C	PSEG03	E21	03/08/2024	1010B
P1724-07	1653	Water	Flash Point	Cool 4 deg C	PSEG03	E21	03/08/2024	1010B

Date/Time 03/11/2024 11:00
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: Jd Woe

Date/Time 03/11/2024 14:30
Raw Sample Received by: Jd Woe
Raw Sample Relinquished by: N.P.W.C.